CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

184

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 04-002516 (#12) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 7/12/05 Cal 1 0.081 AC Cal 2 0.082 AC

Date: 8/4/05 Cal 1 0.078 AC Cal 2 0.079 AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 7/12/05 and 8/4/05 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Irene Doran

Irene Doran

8/8/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 8, day of August, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Terry Gallegos 28847
(Print Name)

INTOXILYZER SERIAL # 64-00 2516 (#12) LOCATION Traffic

DATE August 4 ~~19~~ 2005 TIME 1440

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .339 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.080 AC
Results: 0.078AC and 0.079 AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: _____

SIGNATURE Terry Gallegos

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
08/04/05

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:40
SUBJECT TEST	.000	14:40
AIR BLANK	.000	14:40

DVM= .339 #12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

SN 64-002516 08/04/05
E47.02 14:42
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

#12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6 TSA

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Tucson Police Department Crime Laboratory

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
08/04/05

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:46
CAL. CHECK	.078	14:46
AIR BLANK	.000	14:47

#12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
08/04/05

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:47
CAL. CHECK	.079	14:48
AIR BLANK	.000	14:48

#12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 6
154

4 of 6
P4

SN 64-002516
E47.02
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

08/04/05
14:53

#12

SUBJECT'S NAME

Traffic
INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 6
TSG

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 8-4 2005 Time 1440
Intoxilyzer Serial # 64-002516 (#12) Location Traffic
QA Specialist Terry Gallegos 28847
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.078AC and 0.079 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Terry Gallegos

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
08/04/05

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:49
INVALID SAMPLE	.XXX	14:49
AIR BLANK	.000	14:49

#12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
08/04/05

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:50
*SUBJECT TEST	.088	14:52
AIR BLANK	.000	14:52

* DEFICIENT SAMPLE - VALUE
PRINTED WAS HIGHEST OBTAINED.

#12

SUBJECT'S NAME

Traffic

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

6 of 6
TSG

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.