



TUCSON POLICE DEPARTMENT
CRIME LABORATORY
INTOXILYZER SERVICE RECORD

Date: 01.05.04
Intoxilyzer Serial Number: #13 64-002659
QAS: Terry Gallegos 28847

Problem Description:

Change date to 2004.

If the instrument is taken out of service, complete the following:

Date out of service _____

Calibration Check completed prior to removal from service:

Yes Date: _____ No

SQAP completed prior to removal from service:

Yes Date: _____ No

Date returned to service: _____

Corrective Action Taken or Repairs:

Opened side panel and rotated year dial to "04."