186 LS

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* TSH

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331
(Print Name)

INTOXILYZER SERIAL # 64-002659(#13) LOCATION LAB
DATE Oct 7 19 2004 LS TIME 1405

LS Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

LS DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .354.

LS Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

LS Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

LS Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

LS Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

LS Alcohol free subject Test result 0.000 AC

LS Error recognition logic system functioning
Invalid test printed

LS Proper sample recognition system
Invalid sample printed LS
Deficient sample printed LS

N/A Completeness of sample purge with collection tube**

LS Calibration standard 0.080 AC
Results: 0.078AC & 0.077AC ✓

Instrument operating properly and accurately. YES__ NO__

COMMENTS: _____

SIGNATURE Irene Doran

o N/A LS alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 64-002659 #13 10/07/04
E47.02 14:16
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002659
10/07/04 #13

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:13
SUBJECT TEST	.000	14:14
AIR BLANK	.000	14:14

DVM=.354

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002659
10/07/04

#13

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:06
CAL. CHECK	.078	14:06
AIR BLANK	.000	14:07

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED

INSTRUMENT LOCATION

463311

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002659
10/07/04

#13

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:07
CAL. CHECK	.077	14:08
AIR BLANK	.000	14:08

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED

INSTRUMENT LOCATION

463311

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

396

406
OK

SN 64-002659
E47.02
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

10/07/04
14:21

#13

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

5 8 6 ps

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE 154

Agency Tucson Police Department Date 10/7/04 Time 1405
Intoxilyzer Serial# 04-002659 (#13) Location LAB
QA Specialist Irene Doran (46331)
(Print Name)

(ps) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(ps) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(ps) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(ps) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

ps (ps) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

ps (ps) 6. Push Start Test button or press enter on keyboard.

(ps) 7. Insert card in response to display.

ps (ps) 8. Air blank completed.

ps (ps) 9. Calibration check completed.
Test results 0.078 + 0.077 AC

ps (ps) 10. Air blank completed.

ps (ps) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(ps) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002659
10/07/04 #13

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:19
INVALID SAMPLE	.XXX	14:20
AIR BLANK	.000	14:20

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

4633
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002659
10/07/04 ~~#16~~ #13

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:09
*SUBJECT TEST	.070	14:11
AIR BLANK	.000	14:11

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

696