186
ls

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* D4

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)

INTOXILYZER SERIAL # 64-003303 (#16) LOCATION LAB
DATE Oct 25 ~~18~~ 2004 ls TIME 0900

ls Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

ls DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .573.

ls Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

ls Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

ls * Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

ls Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

Alcohol free subject Test result 0.000 AC

ls Error recognition logic system functioning
Invalid test printed

ls Proper sample recognition system
Invalid sample printed ls
Deficient sample printed ls

N/A Completeness of sample purge with collection tube**

ls Calibration standard 0.080 AC
Results: 0.081AC + 0.081AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: * moved ahead 3min (0914-0917)

SIGNATURE Irene Doran

0.N/A ls alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 64-003303 #16 10/25/04
09:12
E47.04
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |
Q306 |
Q305 |
Q302 |
Q303 |
Q304 |
Q301 |

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
10/25/04 #16

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:31
SUBJECT TEST	.000	09:31
AIR BLANK	.000	09:31

DVM=.573

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

286

#16

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
10/25/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:33
CAL. CHECK	.081	09:33
AIR BLANK	.000	09:33

SUBJECT'S NAME

LAB
TIME FIRST OBSERVED INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

PS

#16

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
10/25/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:34
CAL. CHECK	.081	09:34
AIR BLANK	.000	09:35

SUBJECT'S NAME

LAB
TIME FIRST OBSERVED INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

PS

Tucson Police Department Crime Laboratory

386
PS

486

SN 64-003303 10/25/04
E47.04 09:36
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

#16

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC

5 of 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 10/25/04 Time 0900
Intoxilyzer Serial# 64-003303 (#16) Location LAB
QA Specialist Irene Doran (46331)
(Print Name)

(IDR) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(IDR) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(IDR) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(IDR) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(IDR) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(IDR) 6. Push Start Test button or press enter on keyboard.

(IDR) 7. Insert card in response to display.

(IDR) 8. Air blank completed.

(IDR) 9. Calibration check completed.
Test results 0.081 + 0.081 AC

(IDR) 10. Air blank completed.

(IDR) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(IDR) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

#16

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
10/25/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:40
INVALID SAMPLE	.XXX	09:40
AIR BLANK	.000	09:41

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
10/25/04

#16

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:36
*SUBJECT TEST	.071	09:38
AIR BLANK	.000	09:39

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.