CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 64-003303 (#16) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 10·6·2006 Cal 1 0.080 AC Cal 2 0.081 AC

Date: 11·3·2006 Cal 1 0.076 AC Cal 2 0.079 AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 10·6·2006 and 11·3·2006 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Terry Gallegos

Terry S. Gallegos

12-5-2006

ADPS QAS Permit # QAS 976

SUBSCRIBED AND SWORN to before me this 5th, day of December, 2006.

Rosa Reyna

Notary Public

OFFICIAL SEAL
ROSA REYNA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Jan. 31, 2007

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-4

1 of 6
P4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallego 28847 AGENCY Tucson Police Department
DATE 11-3-2006 TIME 1737
INTOXILYZER SERIAL # 64-003303 (#16) LOCATION DUI HQ

( ✓ ) 1. Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

( ✓ ) 1. DVM Test check. Setting should be between .010 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .372.
( ✓ ) 2. Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".
( ✓ ) 3. Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".
( ✓ ) 4. Clock time check. Mode selection switch S10 on or keyboard menu selection "E".
( ✓ ) 5. Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject Test result 0.000 AC
( ✓ ) 2. Error recognition logic system functioning
Invalid test printed
( ✓ ) 3. Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓
( ✓ ) 4. Standard Calibration Check standard 0.080 AC
Results: 0.076AC and 0.079 AC

Instrument operating accurately and properly. YES ✓ NO_____

COMMENTS: _____

SIGNATURE Terry Gallego

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-4 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 64-003303 11/03/06
17:39
E47.04
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

#16
SUBJECT'S NAME

DUI HQ
TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
11/03/06

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 17:36 |
| SUBJECT TEST | .000 | 17:37 |
| AIR BLANK | .000 | 17:37 |

DVM=.372

#16
SUBJECT'S NAME

DUI HQ
TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
11/03/06

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 18:20 |
| CAL. CHECK | .079 | 18:21 |
| AIR BLANK | .000 | 18:21 |

#16

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Gallegos 28947

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
11/03/06

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 18:19 |
| CAL. CHECK | .076 | 18:19 |
| AIR BLANK | .000 | 18:19 |

#16

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Gallegos 28947

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4of4
D4

SN 64-003303 11/03/06
E47.04 17:49
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

#16

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-2

5 of 6
T54

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Terry Gallegos 28947 AGENCY Tucson Police Department
DATE 11-3-2006 TIME 1737
INTOXILYZER SERIAL # 64-003303 (#16) LOCATION DUI HQ

( ✓ ) 1. Pour a standard alcohol concentration solution into a clean dry simulator jar and assemble the simulator. Ensure that a tight seal has been made.
Standard value: 0.080 AC.

( ✓ ) 2. Turn on the simulator and allow the temperature to reach 34°C ± 0.2°C.

( ✓ ) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

( ✓ ) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

( ✓ ) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

( ✓ ) 6. Push Start Test button or press enter on keyboard.

( ✓ ) 7. Insert card in response to display.

( ✓ ) 8. Air blank completed.

( ✓ ) 9. Standard Calibration Check completed.
Test results 0.076AC and 0.079 AC.

( ✓ ) 10. Air blank completed.

( ✓ ) 11. When test complete, remove printed record. Attach the record to the completed checklist.

( ✓ ) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Terry Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-2 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
11/03/06

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 17:43 |
| INVALID SAMPLE | .XXX | 17:44 |
| AIR BLANK | .000 | 17:44 |

#16

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003303
11/03/06

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 17:45 |
| *SUBJECT TEST | .133 | 17:47 |
| AIR BLANK | .000 | 17:48 |

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

#16

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.