EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)
INTOXILYZER SERIAL # (15) 04-003336 LOCATION Team IV ODE
DATE February 19 ~~19~~ 2004 TIME 0954

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .097.
✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".
✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".
✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".
✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC
✓ Error recognition logic system functioning
✓ Invalid test printed
✓ Proper sample recognition system
✓ Invalid sample printed
✓ Deficient sample printed
N/A Completeness of sample purge with collection tube**
✓ Calibration standard 0.080 AC
Results: 0.078AC + 0.078AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN [illegible] #15 [illegible]
E47.00 [illegible]

[illegible]

SUBJECT'S NAME

CDE

TIME FIRST OBSERVED INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER [illegible]
AZ [illegible] [illegible]
MAR 19, 86 [illegible]

#15

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:55 |
| SUBJECT TEST | .000 | 09:55 |
| AIR BLANK | .000 | 09:56 |

DUMF .297

SUBJECT'S NAME

CDE

TIME FIRST OBSERVED INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

286

CMI INC
INTOXILYZER ALCOHOL ANALYZER
AZ MODEL 5000 SN 64 003336
02/19/04

#15

| TEST | BAL VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:56 |
| CAL. CHECK | .078 | 09:58 |
| AIR BLANK | .000 | 09:58 |

SUBJECT'S NAME

TIME FIRST OBSERVED

CDE
INSTRUMENT LOCATION

DCR AN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

#15

CMI INC
INTOXILYZER ALCOHOL ANALYZER
AZ MODEL 5000 SN 64 003336
02/19/04

| TEST | BAL VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:00 |
| CAL. CHECK | .078 | 10:01 |
| AIR BLANK | .000 | 10:01 |

SUBJECT'S NAME

TIME FIRST OBSERVED

CDE
INSTRUMENT LOCATION

DCR AN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

386

4 of 6

SN 64-003336 02/19/04
E47.04 10:10
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

TIME FIRST OBSERVED

ODE
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

5 8 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 2/19/04 Time 0954
Intoxilyzer Serial# (15)64-003336 Location Team IV ODE
QA Specialist Irene Doran 46331
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.078 + 0.078 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003336
02/19/04

#15

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:05 |
| INVALID SAMPLE | .XXX | 10:05 |
| AIR BLANK | .000 | 10:06 |

DORAN
SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN 46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Binaca

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-003336
02/19/04

#15

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:13 |
| *SUBJECT TEST | .078 | 10:15 |
| AIR BLANK | .000 | 10:15 |

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sample did not meet time requirement.

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.