CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-004014 (#17) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 9·8·2006 Cal 1 0.078 AC Cal 2 0.081 AC

Date: 10·6·2006 Cal 1 0.080 AC Cal 2 0.081 AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 9·8·2006 and 10·6·2006 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Terry Gallegos

Terry S. Gallegos

10·30·2006

ADPS QAS Permit # QAS 976

SUBSCRIBED AND SWORN to before me this 30, day of October, 2006.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 6
TSG

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 10-6-2006 TIME 1652

INTOXILYZER SERIAL # 66-004014 (#17) LOCATION DUI HQ

(✓) 1. Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

(✓) 1. DVM Test check. Setting should be between .010 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .425.

(✓) 2. Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

(✓) 3. Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

(✓) 4. Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

(✓) 5. Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject Test result 0.000 AC

(✓) 2. Error recognition logic system functioning
Invalid test printed

(✓) 3. Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

(✓) 4. Standard Calibration Check standard 0.080 AC
Results: 0.080AC and 0.081 AC

Instrument operating accurately and properly. YES ✓ NO ____

COMMENTS: _____

SIGNATURE Terry Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-4 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-004014 10/06/06
E647.05 16:59
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

#17

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
10/06/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:52
SUBJECT TEST	.000	16:53
AIR BLANK	.000	16:53

DVM=.425 #17

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Tucson Police Department Crime Laboratory

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
10/06/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:56
CAL. CHECK	.081	16:56
AIR BLANK	.000	16:57

#17

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
10/06/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	16:54
CAL. CHECK	.080	16:55
AIR BLANK	.000	16:55

#17

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 6
pg 104

SN 66-004014
E647.05
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

10/06/06
17:06

#17

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

5 of 6
TSG

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 10-6-2006 TIME 1652
INTOXILYZER SERIAL # 66-004014 (#17) LOCATION DUI HQ

(✓) 1. Pour a standard alcohol concentration solution into a clean dry simulator jar and assemble the simulator. Ensure that a tight seal has been made. Standard value: 0.080 AC.
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C ± 0.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Standard Calibration Check completed. Test results 0.080AC and 0.081 AC.
(✓) 10. Air blank completed.
(✓) 11. When test complete, remove printed record. Attach the record to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Terry Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-2 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
10/06/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	17:03
*SUBJECT TEST	.064	17:05
AIR BLANK	.000	17:05

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

#17

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
10/06/06

TEST	BAC VALUE	TIME
AIR BLANK	.000	17:01
INVALID SAMPLE	.XXX	17:02
AIR BLANK	.000	17:02

#17

SUBJECT'S NAME

DUI HQ

INSTRUMENT LOCATION

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.