CERTIFICATION by CUSTODIAN of RECORDS BM

*Tucson Police Department Crime Laboratory*

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-001759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 7/13/06 Cal 1 0.078 AC Cal 2 0.078 AC

Date: 8/11/06 Cal 1 0.079 AC Cal 2 0.080 AC

Date: N/A Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 7/13/06 and 8/11/06 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Jennifer L. Coombs

Jennifer L. Coombs

8/11/06

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 11, day of August, 2006.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD QUALITY ASSURANCE PROCEDURE 54

1 of 4 file

QA SPECIALIST Jennifer Coombs 46160 AGENCY Tucson Police Department
DATE August 11, 2006 TIME 1043
INTOXILYZER SERIAL # 66-001759 LOCATION DUI HTD

( ✓ ) 1. Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

( ✓ ) 1. DVM Test check. Setting should be between .010 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .394.
( ✓ ) 2. Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".
( ✓ ) 3. Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".
( ✓ ) 4. Clock time check. Mode selection switch S10 on or keyboard menu selection "E".
( ✓ ) 5. Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject Test result 0.000 AC
( ✓ ) 2. Error recognition logic system functioning
Invalid test printed
( ✓ ) 3. Proper sample recognition system
Invalid sample printed
Deficient sample printed
( ✓ ) 4. Standard Calibration Check standard 0.080 AC
Results: 0.079AC ≠ 0.080 AC

Instrument operating accurately and properly. YES ✓ NO ____

COMMENTS: _____

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-4 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

```
SN 66-001759                    08/11/06
E767.14                            10:45
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |-------------------------
Q306 |-------------------------
Q305 |-------------------------
Q302 |-------------------------
Q303 |-------------------------
Q304 |-------------------------
Q301 |-------------------------
```

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

```
TUCSON POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000        SN 66-001759
08/11/06                     10:47

TEST            BRAC VALUE    TIME
AIR BLANK          .000       10:47
SUBJECT TEST       .000       10:47
AIR BLANK          .000       10:47
```

DVM=.394

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

# Tucson Police Department Crime Laboratory

TUCSON POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
08/11/06 10:55

| TEST | BRAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:55 |
| CAL. CHECK | .079 | 10:55 |
| AIR BLANK | .000 | 10:55 |

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

TUCSON POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
08/11/06 10:57

| TEST | BRAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:56 |
| CAL. CHECK | .080 | 10:57 |
| AIR BLANK | .000 | 10:57 |

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 6

4 of 6
gh
BY

SN 66-001759
E767.14
INVALID TEST

08/11/06
10:48

PRESSED START TEST AT
IMPROPER TIME

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT E-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

5 of 6 JL

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE DH

QA SPECIALIST Jennifer Coombs 46160 AGENCY Tucson Police Department
DATE August 11, 2006 TIME 1043
INTOXILYZER SERIAL # 66-001759 LOCATION DUI HQ

( ✓ ) 1. Pour a standard alcohol concentration solution into a clean dry simulator jar and assemble the simulator. Ensure that a tight seal has been made. Standard value: 0.080 AC.
( ✓ ) 2. Turn on the simulator and allow the temperature to reach 34°C ± 0.2°C.
( ✓ ) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
( ✓ ) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
( ✓ ) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
( ✓ ) 6. Push Start Test button or press enter on keyboard.
( ✓ ) 7. Insert card in response to display.
( ✓ ) 8. Air blank completed.
( ✓ ) 9. Standard Calibration Check completed. Test results 0.079 AC & 0.080 AC.
( ✓ ) 10. Air blank completed.
( ✓ ) 11. When test complete, remove printed record. Attach the record to the completed checklist.
( ✓ ) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-2 (Rev. 05-1)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
08/11/06 10:49

| TEST | BRAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:49 |
| INVALID SAMPLE | .MOA | 10:49 |
| AIR BLANK | .000 | 10:49 |

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
08/11/06 10:53

| TEST | BRAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:50 |
| *SUBJECT TEST | .230 | 10:53 |
| AIR BLANK | .000 | 10:53 |

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 6