CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

104

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-003327 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 6/14/05 Cal 1 0.079 AC Cal 2 0.078 AC

Date: 6/27/05 Cal 1 0.080 AC Cal 2 0.079 AC

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 6/14/05 and 6/27/05 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on NA.
The instrument was returned to service on NA.

Irene Doran

Irene Doran

7/6/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 6, day of July, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Terry Gallegos 28847
(Print Name)

INTOXILYZER SERIAL # 66-003327 LOCATION Lab

DATE June 27 ~~19~~ TSH 2005 TIME 1200

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .278 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.080 AC
Results: 0.080AC and 0.079 AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: _____

SIGNATURE Terry Gallegos

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-003327 06/27/05
E767.14 12:03
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

Lab

TIME FIRST OBSERVED INSTRUMENT LOCATION

28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
06/27/05 12:00

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:59
SUBJECT TEST	.000	12:00
AIR BLANK	.000	12:00

DVM = .278

SUBJECT'S NAME

Lab

TIME FIRST OBSERVED INSTRUMENT LOCATION

28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Tucson Police Department Crime Laboratory

3 of 7

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
06/27/05 12:05

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:04
CAL. CHECK	.080	12:05
AIR BLANK	.000	12:05

SUBJECT'S NAME

TIME FIRST OBSERVED

Lab
INSTRUMENT LOCATION

28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
06/27/05 12:08

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:07
CAL. CHECK	.079	12:08
AIR BLANK	.000	12:08

SUBJECT'S NAME

TIME FIRST OBSERVED

Lab
INSTRUMENT LOCATION

28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-003327 06/27/05
E767.14 12:10
INVALID TEST

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

Lab
TIME FIRST OBSERVED INSTRUMENT LOCATION

28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7
TDH

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 6·27·2005 Time 1200
Intoxilyzer Serial # 66-003327 Location Lab
QA Specialist Terry Gallegos 28847
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓✓) 6. Push Start Test button or press enter on keyboard.
(✓✓) 7. Insert card in response to display.
(✓✓) 8. Air blank completed.
(✓✓) 9. Calibration check completed.
Test results 0.080 AC and 0.079 AC
(✓✓) 10. Air blank completed.
(✓✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Terry Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
06/27/05 - 12:11

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:10
INVALID SAMPLE	.MOA	12:11
AIR BLANK	.000/	12:11

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

Lab
INSTRUMENT LOCATION

28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
06/27/05 12:16

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:12
*SUBJECT TEST	.086	12:15
AIR BLANK	.000	12:16

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

Lab
INSTRUMENT LOCATION

28847
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

6 of 7
Pg

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

Date: 6·27·2005

QAS: Terry Gallegos 28847

Intoxilyzer Serial Number: 66-003327

Reason For Service: O'Hara called to say "printer error" on instrument 6·24·05. Brought to Lab 6·27·05 by Sgt. Leotaud.

Date Out of Service: N/A

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action: ☐ Sent to Applied Electronics

Opened Intox and removed small piece of printcard in the printer gear. Cleaned roller bar and realigned printer head. Diagnostics ok. Performed QAS and returned Intox to DUI HQ 6·27·2005

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.