CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

154

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 606-003327 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 1/26/06 Cal 1 0.078 AC Cal 2 0.079 AC

Date: 2/23/06 Cal 1 0.077 AC Cal 2 0.080 AC

Date: _____ Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 1/26/06 and 2/23/06 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Jennifer L. Coombs

Jennifer L. Coombs

3/8/06

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 8, day of March, 2006.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A. Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 154
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 46160
(Print Name)

INTOXILYZER SERIAL # 66-003327 LOCATION DUI HQ
DATE February 23 ~~19~~ 2006 TIME 1544

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .400 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓* Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓✓ Calibration standard 0.080 AC AC
Results: 0.077 AC & 0.080 AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: * moved clock 1547 → 1553

SIGNATURE Jennifer L. Coombs

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

```
SN 66-003327                02/23/06,
E767.14                        15:54,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH,
Q307 |--------------------------,
Q306 |--------------------------,
Q305 |--------------------------,
Q302 |--------------------------,
Q303 |--------------------------,
Q304 |--------------------------,
Q301 |--------------------------,
```

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

```
TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER ,
AZ MODEL 5000        SN 66-003327
02/23/06                    15:58 ,

TEST            BRAC VALUE    TIME
AIR BLANK          .000       15:57
SUBJECT TEST       .000       15:57
AIR BLANK          .000       15:58
```

DYM = .400

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED

INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Tucson Police Department Crime Laboratory

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
02/23/06 16:02

TEST	BRAC VALUE	TIME
AIR BLANK	.000	16:02
CAL. CHECK	.077	16:02
AIR BLANK	.000	16:02

SUBJECT'S NAME

TIME FIRST OBSERVED

DWI HQ
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-003327
02/23/06 16:05

TEST	BRAC VALUE	TIME
AIR BLANK	.000	16:04
CAL. CHECK	.080	16:04
AIR BLANK	.000	16:05

SUBJECT'S NAME

TIME FIRST OBSERVED

DWI HQ
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 6

4 of 6
file

SN 66-003327
E767.14
INVALID TEST

02/23/06,
15:59,

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 6 file

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

134

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 2/23/06 Time 1544
Intoxilyzer Serial # 66-003327 Location DUI HQ
QA Specialist Jennifer Coombs 46160
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
✓ (✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
✓ (✓) 6. Push Start Test button or press enter on keyboard.
✓ (✓) 7. Insert card in response to display.
✓ (✓) 8. Air blank completed.
✓ (✓) 9. Calibration check completed.
Test results 0.077AC & 0.080 AC
✓ (✓) 10. Air blank completed.
✓ (✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER ,
AZ MODEL 5000 SN 66-003327
02/23/06 16:11 ,

TEST	BRAC VALUE	TIME
AIR BLANK	.000	16:10
INVALID SAMPLE	.MOA	16:11
AIR BLANK	.000	16:11

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER ,
AZ MODEL 5000 SN 66-003327
02/23/06 16:21 ,

TEST	BRAC VALUE	TIME
AIR BLANK	.000	16:17
*SUBJECT TEST	.189	16:20
AIR BLANK	.000	16:21

* DEFICIENT SAMPLE - VALUE
PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 6