106
PS

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 184
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331
(Print Name)

INTOXILYZER SERIAL #66-004014 (#17) LOCATION DUI HQ
DATE November 4 ~~18~~ 2004 TIME 1030
PS

PS Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

PS DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .396.

PS Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

PS Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

PS Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

PS Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

PS Alcohol free subject Test result 0.000 AC

PS Error recognition logic system functioning
Invalid test printed

PS Proper sample recognition system
Invalid sample printed PS
Deficient sample printed PS

N/A Completeness of sample purge with collection tube**

PS Calibration standard 0.080 AC
Results: 0.077AC & 0.078AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

Q. N/A PS alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 66-004014 #17 11/04/04
E647.05 11:09
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

TIME FIRST OBSERVED DUI HQ INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

PA

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
11/04/04 #17

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:15
SUBJECT TEST	.000	11:15
AIR BLANK	.000	11:16

DVM = .396

SUBJECT'S NAME

TIME FIRST OBSERVED DUI HQ INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

PA

286

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
11/04/04 #17

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:37
CAL. CHECK	.077	11:37
AIR BLANK	.000	11:38

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
11/04/04 #17

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:38
CAL. CHECK	.078	11:39
AIR BLANK	.000	11:39

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

386

4 of 6

SN 66-0042.- 11/04/04
E647.05 11:14
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

5 & 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 11/4/04 Time 1030
Intoxilyzer Serial# 66-004014 (#17) Location DUI HQ
QA Specialist Irene Doran 46331
(Print Name)

(ID) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(ID) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(ID) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(ID) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(ID) (ID) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(ID) (ID) 6. Push Start Test button or press enter on keyboard.

(ID) (ID) 7. Insert card in response to display.

(ID) (ID) 8. Air blank completed.

(ID) (ID) 9. Calibration check completed.
Test results 0.077 + 0.078 AC

(ID) (ID) 10. Air blank completed.

(ID) (ID) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(ID) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
11/04/04

#17

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:40
INVALID SAMPLE	.XXX	11:41
AIR BLANK	.000	11:41

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004014
11/04/04

#17

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:42
*SUBJECT TEST	.223	11:44
AIR BLANK	.000	11:45

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.