CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

184

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-004015 (#18) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 4/20/05 Cal 1 0.081 AC Cal 2 0.082 AC

Date: 5/18/05 Cal 1 0.079 AC Cal 2 0.079 AC

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 4/20/05 and 5/18/05 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on NA.
The instrument was returned to service on NA.

Irene Doran

Irene Doran

5/19/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 19, day of May, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1076

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331
(Print Name)

INTOXILYZER SERIAL # 66-004015 (#18) LOCATION DUI HQ

DATE May 18 2005 TIME 1028

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .383 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓✓ Calibration standard 0.080 AC
Results: 0.079AC + 0.079AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-004015 #18 05/18/05
E647.05 10:29
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
05/18/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:30
SUBJECT TEST	.000	10:31
AIR BLANK	.000	10:32

DVM=.383

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

386

Tucson Police Department Crime Laboratory

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
05/18/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:46
CAL. CHECK	.079	10:46
AIR BLANK	.000	10:46

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
05/18/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:44
CAL. CHECK	.079	10:45
AIR BLANK	.000	10:45

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

486
P5

SN 66-004015
E647.05
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

05/18/05
10:43

#18

SUBJECT'S NAME

DUI HQ
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

CMI

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5876
P5

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE P4

Agency Tucson Police Department Date 5/18/05 Time 1028
Intoxilyzer Serial # 66-004015 (#18) Location DUI HQ
QA Specialist Irene Doran #46331
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.079 + 0.079 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
05/18/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:38
INVALID SAMPLE	.XXX	10:39
AIR BLANK	.000	10:39

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS: sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
05/18/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:40
*SUBJECT TEST	.232	10:42
AIR BLANK	.000	10:42

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.