CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-004015(#18) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 8/8/05 Cal 1 0.078 AC Cal 2 0.077 AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Between the dates of — and — the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 7/7/05.
The instrument was returned to service on 8/8/05.

Irene Doran

Irene Doran

8/9/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 9, day of August, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

MEMORANDUM

DATE: July 13, 2005

TO: Applied Electronics
448 Grand Avenue
Eagle, CO 81631

FROM: **Irene Doran**
Tucson Police Department
Crime Laboratory
270 S. Stone Avenue
Tucson, AZ 85701-1917

SUBJECT: Intoxilyzer 5000 Maintenance and Repair

Enclosed is Intoxilyzer 66-004015 (#18). The print is very light on the printcards.

Please repair as needed.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CUSTOMER SERVICE REPORT

RECEIVED FROM	
CUSTOMER	Tucson Police Dept
DEPT.	Crime Lab
STREET	270 S. Stone Ave
CITY	Tucson
STATE	AZ ZIP 85701
ATTEN:	
PHONE ( )	

DATE	7-27-05
REC'D VIA	UPS
COD CHARGE	
P.O. NO.	

SHIP TO	
CUSTOMER	
DEPT.	
STREET	
CITY	
STATE	
ATTEN:	
PHONE ( )	

Pursuant to section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EQUIPMENT RECEIVED INVENTORY

MODEL	5000	CARRYING CASE	☐	EXTENSION CABLE	☐	KEYBOARD/CABLE	None ☐
SERIAL NO.	66-004015	ECM CABLE	☐	TUNING FORK POUCH	☐	PAT BOX	☐
ANTENNA SER.NO.		REMOTE LOCK	☐	TUNING FORK POUCH	☐	SIMULATOR	☐
ANTENNA SER.NO.		DASH BRACKET	☐	BATTERY PACK	☐	BREATH TUBE	☒
E.C.M./VM SER.NO.		DECK BRACKET	☐	BATTERY CHARGER	☐		☐
TUNING FORK SER.NO.		BAIL BRACKET	☐	A/C POWER CORD	None ☐		☐
TUNING FORK SER.NO.							
MISC.SER.NO.							

DESCRIPTION OF CHARGES	PART NO.	QTY	UNIT PRICE	TOTAL
Breath hose Cover	800075	1		
Calibration Cert		1		

TOTAL PARTS	
2.0 HRS TOTAL LABOR	
TAX	
45 LBS INSURANCE/SHIPPING/HANDLING	
TOTAL	

CUSTOMER COMPLAINT / SERVICE REQUESTED

Prints out very light

CORRECTIVE ACTION TAKEN

Repaired loose print head. Cleaned adjusted printer for best quality. Replaced breath hose cover. Cleaned optics. Certified Calibration

FACTORY AUTHORIZED WARRANTY
SERVICE FACILITY FOR ALL
CMI AND MPH PRODUCTS

APPLIED ELECTRONICS, INC.
448 GRAND AVE. • P.O. BOX 658 • EAGLE, CO 81631 • Phone: (970) 328-5420

Pursuant to section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

APPLIED ELECTRONICS Inc.

~~55 EAGLE PARK EAST DRIVE~~ 1445 GRAND AVE. • P.O. BOX 638 • EAGLE, CO 81631 Phone: (970) 328-5420

STATEMENT OF SERVICE WARRANTY

Applied Electronics warrants that any component having been replaced, to be free of defects in material and workmanship for a period of 90 days from date of delivery to its customer. This warranty is limited to only components that have been replaced during the repair of your product. Applied Electronics obligation is limited to repairing or replacing as we may elect of any component covered under this service warranty.

Any parts or components considered covered by this warranty shall be returned freight prepaid to Applied Electronics. Return freight shall be paid by Applied Electronics.

This warrnaty may be voided if the product has been adversely effected or tampered with in any way.

This service warranty is separate from the original manufacturers warranty. In no event shall Applied Electronics be liable for any loss or damages arising out of any such defective component or workmanship.

CERTIFICATE OF CALIBRATION

Pursuant to section 28 1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

This is to certify that the calibration of INTOXILYZER®
serial number 66-004015, *was tested and found to be in compliance*
with the National Highway Traffic Safety Administration Standard
for Devices to Measure Breath Alcohol
(3.8 F.R., No. 212 30459-30463 Nov. 5, 1973.)

Date Aug 2, 2005 Signed Robert C Bank
Technician

APPLIED ELECTRONICS

Factory Authorized Service Facility For CMI/MPH Products

1 of 1

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

TSH

QAS: Irene Doran 46331

Intoxilyzer Serial Number: 66-004015 (#18)

Date Returned to Service: 8/8/05

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1076

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST 46331 Irene Doran
(Print Name)

INTOXILYZER SERIAL # 66-004015 (#18) LOCATION LAB

DATE August 8 ~~19~~ 2005 TIME 1415

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .369 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓✓ Calibration standard 0.080 AC
Results: 0.078AC + 0.077AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-004015 #18 08/08/05
E647.05 14:17
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

46331 2
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
08/08/05 #18

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:15
SUBJECT TEST	.000	14:16
AIR BLANK	.000	14:16

DVM=1.369

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Tucson Police Department Crime Laboratory

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
08/08/05

#18

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:22
CAL. CHECK	.078	14:23
AIR BLANK	.000	14:23

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
08/08/05

#18

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:24
CAL. CHECK	.077	14:24
AIR BLANK	.000	14:24

SUBJECT'S NAME

LAB
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

376
PS
BH

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 6

SN 66-004015 08/08/05
E647.05 14:20
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

#18

LAB

SUBJECT'S NAME

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

586

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 8/8/05 Time 1415
Intoxilyzer Serial # 66-004015 (#18) Location LAB
QA Specialist Irene Doran 46331
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.078 + 0.077 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
08/08/05

#18

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:35
INVALID SAMPLE	.XXX	14:36
AIR BLANK	.000	14:37

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004015
08/08/05

#18

TEST	BAC VALUE	TIME
AIR BLANK	.000	14:37
*SUBJECT TEST	.192	14:39
AIR BLANK	.000	14:40

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.