106

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)

INTOXILYZER SERIAL # 06-004798 LOCATION ~~ODE~~ ODW

DATE March 14 ~~19~~ 2004 TIME 1528

Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .367.

Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

Alcohol free subject Test result 0.000 AC

Error recognition logic system functioning
Invalid test printed

Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

Calibration standard 0.080 AC
Results: 0.074 AC 0.076 AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0.N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

E767.14 15:39



SUBJECT'S NAME

TIME FIRST OBSERVED

ODW

INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
03/14/04 15:30

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:29
SUBJECT TEST	.000	15:30
AIR BLANK	.000	15:30

DVM = .367

SUBJECT'S NAME

TIME FIRST OBSERVED

ODW

INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

286

AZ MODEL 5000 SN 66-004798
03/14/04 15:31

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:31
CAL. CHECK	.074	15:31
AIR BLANK	.000	15:31

SUBJECT'S NAME

TIME FIRST OBSERVED

ODW
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
03/14/04 15:33

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:32
CAL. CHECK	.075	15:33
AIR BLANK	.000	15:33

SUBJECT'S NAME

TIME FIRST OBSERVED

ODW
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

386

4 8 6

[illegible]
[illegible]
INVALID TEST

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:41

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

TIME FIRST OBSERVED

ODW
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

586
65

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 3/14/04 Time 1528
Intoxilyzer Serial# 66-004798 Location ODW
QA Specialist Irene Doran (46331)
(Print Name)

(ID) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(ID) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(ID) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(ID) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(ID) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(ID) 6. Push Start Test button or press enter on keyboard.

(ID) 7. Insert card in response to display.

(ID) 8. Air blank completed.

(ID) 9. Calibration check completed.
Test results 0.074 & 0.076 AC

(ID) 10. Air blank completed.

(ID) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(ID) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

A2 MODEL 5000 SN 66-004798
03/14/04 15:43

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:42
INVALID SAMPLE	.MOA	15:43
AIR BLANK	.000	15:43

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

ODW

TIME FIRST OBSERVED

INSTRUMENT LOCATION

DORAN 46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Binaca

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

A2 MODEL 5000 SN 66-004798
03/14/04 15:37

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:34
*SUBJECT TEST	.073	15:37
AIR BLANK	.000	15:37

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

ODW

TIME FIRST OBSERVED

INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.