



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

[Handwritten signature]

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-004798 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 5.11.06 Cal 1 0.079 AC Cal 2 0.078 AC

Date: 6.8.06 Cal 1 0.078 AC Cal 2 0.076 AC

Date: Cal 1 AC Cal 2 AC

Between the dates of 5.11.06 and 6.8.06 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on .
The instrument was returned to service on .

Terry S. Gallegos
Terry S. Gallegos

6.13.2006

ADPS QAS Permit # QAS 976

SUBSCRIBED AND SWORN to before me this 13, day of June, 2006.



Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6
Note: The applicable A.A.C. document sections were revised effective May 18, 2006 at 0900 hours. The new sections are A.A.C. R13-10-104.A.5 and A.A.C. R13-10-104.A.2

1 of 6
TS6

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD QUALITY ASSURANCE PROCEDURE

Handwritten initials

QA SPECIALIST Terry S. Gallegos 28847 AGENCY Tucson Police Department
DATE 06-08-2006 TIME 1130
INTOXILYZER SERIAL # 66-004798 LOCATION DDD

(✓) 1. Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

- (✓) 1. DVM Test check. Setting should be between .010 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is 0.329.
- (✓) 2. Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".
- (✓) 3. Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".
- (✓) 4. Clock time check. Mode selection switch S10 on or keyboard menu selection "E".
- (✓) 5. Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject Test result 0.000 AC
- (✓) 2. Error recognition logic system functioning
Invalid test printed
- (✓) 3. Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓
- (✓) 4. Standard Calibration Check standard 0.080 AC
Results: 0.078 AC and 0.076 AC

Instrument operating accurately and properly. YES NO

COMMENTS:

SIGNATURE

Terry S. Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

3 of 6
154
jle

Tucson Police Department Crime Laboratory

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
06/08/06 11:38

TEST BRAC VALUE TIME
AIR BLANK .000 11:38
CAL. CHECK .076 11:38
AIR BLANK .000 11:38

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
06/08/06 11:37

TEST BRAC VALUE TIME
AIR BLANK .000 11:36
CAL. CHECK .078 11:36
AIR BLANK .000 11:37

SUBJECT'S NAME
TIME FIRST OBSERVED
INSTRUMENT LOCATION
OPERATOR

ODD

Gallegos 28947

ADDITIONAL INFORMATION AND/OR REMARKS



INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC

SUBJECT'S NAME
TIME FIRST OBSERVED
INSTRUMENT LOCATION
OPERATOR

ODD

Gallegos 28947

ADDITIONAL INFORMATION AND/OR REMARKS



INTOXILYZER INSTRUMENT PRINTER CARD

© 1986 by CMI INC

4 of 6
754

He

SH 66-804708
2787.14
INVALID TEST

06/08/86
11:45

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME	
TIME FIRST OBSERVED	INSTRUMENT LOCATION
OPERATOR	
ADDITIONAL INFORMATION AND / OR REMARKS	

ODD

Gallegos 28847

INTOXILYZER INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

506
TS4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Terry S. Gallegos 28947 AGENCY Tucson Police Department
DATE 06-08-2006 TIME 1130
INTOXILYZER SERIAL # 66-004798 LOCATION ODD

- (✓) 1. Pour a standard alcohol concentration solution into a clean dry simulator jar and assemble the simulator. Ensure that a tight seal has been made. Standard value: 0.080 AC.
- (✓) 2. Turn on the simulator and allow the temperature to reach 34°C ± 0.2°C.
- (✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
- (✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
- (✓✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
- (✓✓) 6. Push Start Test button or press enter on keyboard.
- (✓✓) 7. Insert card in response to display.
- (✓✓) 8. Air blank completed.
- (✓✓) 9. Standard Calibration Check completed. Test results 0.078AC and 0.076 AC.
- (✓✓) 10. Air blank completed.
- (✓✓) 11. When test complete, remove printed record. Attach the record to the completed checklist.
- (✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE

Terry S. Gallegos

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DPS Form Exh E-2 (Rev. 05-1)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
06/09/06 11:40

TEST BRAC VALUE TIME
AIR BLANK .000 11:40
INVALID SAMPLE .NDA 11:40
AIR BLANK .000 11:40

BEGIN NEW DEPRIVATION PERIOD

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
06/08/06 11:44

TEST BRAC VALUE TIME
AIR BLANK .000 11:44
SUBJECT TEST .166 11:44
AIR BLANK .000 11:44

* DEFICIENT SAMPLE VALUE
PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME
INSTRUMENT LOCATION

ODD

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath



INTOXILYZER* INSTRUMENT PRINTER CARD

© 1986 by CMI INC

SUBJECT'S NAME
INSTRUMENT LOCATION

ODD

TIME FIRST OBSERVED

Gallegos 28847

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS



INTOXILYZER* INSTRUMENT PRINTER CARD

© 1986 by CMI INC

606154 jfe