186

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331
(Print Name)

INTOXILYZER SERIAL # 66-004814 LOCATION LAB

DATE June 9 ~~18~~ 2004 TIME 1237

_____ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

_____ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .375.

_____ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

_____ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

_____ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

_____ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

_____ Alcohol free subject Test result 0.000 AC

_____ Error recognition logic system functioning
Invalid test printed

_____ Proper sample recognition system
Invalid sample printed _____
Deficient sample printed _____

N/A Completeness of sample purge with collection tube**

_____ Calibration standard 0.080 AC
Results: 0.079AC 0.079AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 66-004814 06/09/04
E767.14 12:49



SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004814
06/09/04 12:52

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:51
SUBJECT TEST	.000	12:51
AIR BLANK	.000	12:52

DVM = .375

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004814
06/09/04 12:45

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:44
CAL. CHECK	.079	12:44
AIR BLANK	.000	12:45

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P5

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004814
06/09/04 12:47

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:47
CAL. CHECK	.079	12:47
AIR BLANK	.000	12:47

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P5

Tucson Police Department Crime Laboratory

P4

396
P5

486

SN 66-004814
E767.14
INVALID TEST

06/09/04
12:54

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

596

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 6/9/04 Time 1237
Intoxilyzer Serial# 66-004814 Location LAB
QA Specialist Irene Doran 46331
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.079 + 0.079 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004814
06/09/04 12:53

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:52
INVALID SAMPLE	.MOA	12:53
AIR BLANK	.000	12:53

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Binoca

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004814
06/09/04 12:42

TEST	BRAC VALUE	TIME
AIR BLANK	.000	12:38
*SUBJECT TEST	.070	12:41
AIR BLANK	.000	12:42

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.