1 of 6
jlc

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 46160
(Print Name)

INTOXILYZER SERIAL # 66-004922 LOCATION ODE

DATE August 9 ~~19~~ 2004 jlc TIME 1103

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .422.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.080 AC
Results: 0.079AC & 0.078AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: ___

SIGNATURE Jennifer L. Coombs

0.N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 66-004922 08/09/04
11:09
E767.14)

```
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
0307 |-----------------------------
0306 |-----------------------------
0305 |-----------------------------
0302 |-----------------------------
0303 |-----------------------------
0304 |-----------------------------
0301 |-----------------------------
```

SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004922
08/09/04 11:12

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:11
SUBJECT TEST	.000	11:11
AIR BLANK	.000	11:12

DVM = .422

SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC

file

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004922
08/09/04 11:06

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:05
CAL. CHECK	.079	11:06
AIR BLANK	.000	11:06

SUBJECT'S NAME

TIME FIRST OBSERVED

ODE
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004922
08/09/04 11:08

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:07
CAL. CHECK	.078	11:07
AIR BLANK	.000	11:07

SUBJECT'S NAME

TIME FIRST OBSERVED

ODE
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

3 96

4of6
flc

SN 66-004922
E767.14
INVALID TEST

08/09/04
11:12

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

ODE

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

5 of 6
file

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE DH

Agency Tucson Police Department Date 8/9/04 Time 1103
Intoxilyzer Serial# 166-004922 Location ODE
QA Specialist Jennifer Coombs 46160
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

✓(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

✓(✓) 6. Push Start Test button or press enter on keyboard.

✓(✓) 7. Insert card in response to display.

✓(✓) 8. Air blank completed.

✓(✓) 9. Calibration check completed.
Test results 0.079AC & 0.078 AC

✓(✓) 10. Air blank completed.

✓(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004922
08/09/04 11:14

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:13
INVALID SAMPLE	.MOA	11:14
AIR BLANK	.000	11:14

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46660
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004922
08/09/04 11:18

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:15
*SUBJECT TEST	.067	11:18
AIR BLANK	.000	11:18

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

ODE
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46660
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.