CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

154

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 66-005003 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 10/26/05 Cal 1 0.081 AC Cal 2 0.080 AC

Date: 11/23/05 Cal 1 0.081 AC Cal 2 0.079 AC

Date: — Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 10/26/05 and 11/23/05 the instrument was working properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Irene Doran Johnson

Irene Doran Johnson

12/1/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 1, day of December, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

186 IJ

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 154
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene D. Johnson 4633, (Print Name)

INTOXILYZER SERIAL # 66-005003 LOCATION ODM

DATE November 23 ~~19~~ IJ 2005 TIME 1112

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .387 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ * Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ ✓ Calibration standard 0.080 AC
Results: 0.081AC + 0.079AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: * moved ahead 5minutes

SIGNATURE Irene D. Johnson 4633

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 66-005003 11/23/05
E767.14 11:12
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
11/23/05 11:13

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:13
SUBJECT TEST	.000	11:13
AIR BLANK	.000	11:13

DVM=.387

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

296
IJ

Tucson Police Department Crime Laboratory

396
LJ

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
11/23/05 11:23

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:23
CAL. CHECK	.081	11:23
AIR BLANK	.000	11:23

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
11/23/05 11:25

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:24
CAL. CHECK	.079	11:25
AIR BLANK	.000	11:25

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4086
EJ

SN 66-005003 11/23/05
E767.14 11:14
INVALID TEST

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

586
IJ

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 11/23/05 Time 1112
Intoxilyzer Serial # 66-005003 Location ODM
QA Specialist Irene Johnson 46331
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.081 AC + 0.079 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene D. Johnson #46331

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
11/23/05 11:15

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:15
INVALID SAMPLE	.MOA	11:15
AIR BLANK	.000	11:15

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
11/23/05 11:20

TEST	BRAC VALUE	TIME
AIR BLANK	.000	11:16
*SUBJECT TEST	.239	11:19
AIR BLANK	.000	11:20

* DEFICIENT SAMPLE - VALUE
PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

696
IT