CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

154

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 06-005003 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Between the dates of — and —
the instrument was working properly and accurately.

The instrument was taken out of service for repair on 2/15/06.
The instrument was returned to service on NA.

Irene Doran Johnson

Irene Doran Johnson

2/21/06

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 21, day of February, 2006.

Kathy A Kinney

Notary Public

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

MEMORANDUM

DATE: **2/15/06**

TO: Whom it may concern

FROM: Irene Doran Johnson IJ
Intoxilyzer 5000 QAS
Tucson Police Department
Crime Lab
270 S. Stone Avenue
Tucson, AZ 85701-1917
(520) 791-4494

SUBJECT: Intoxilyzer Out of Service

Intoxilyzer 5000, serial number **66-005003**, was taken out of service for repair on **February 15, 2006**.

I cannot testify that it was working properly and accurately between the dates of **January 19, 2006** and **February 15, 2006**.

This instrument was located at **ODM**.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

184

Date: 2/15/06

QAS: Irene D. Johnson

Intoxilyzer Serial Number: 66-005003

Reason For Service:

Unable to complete monthly checks. Instrument had several errors and finally a Processor Error 1.

Date Out of Service: 2/15/06

Calibration Check completed prior to removal from service

Date: NA

SQAP completed prior to removal from service

Date: NA

Corrective Action: ☑ Sent to Applied Electronics

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

104 EJ

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 84
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene D. Johnson 4633
(Print Name)

INTOXILYZER SERIAL # 66-005003 LOCATION ODM

DATE February 15 ~~18~~ EJ 2006 TIME 1018

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

— DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is unstable

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

— Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

— Calibration standard 0.080 AC

— Results: 0.

Instrument operating properly and accurately. YES ___ NO ✓

COMMENTS: _____

SIGNATURE Irene D. Johnson 4633

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

```
SN 66-005003                    02/15/06
E767.14                            10:21
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |---------------------------
Q306 |---------------------------
Q305 |---------------------------
Q302 |---------------------------
Q303 |---------------------------
Q304 |---------------------------
Q301 |---------------------------
```

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

```
TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000          SN 66-005003
02/15/06                      10:20

TEST              BRAC VALUE    TIME
AIR BLANK            .000       10:19
SUBJECT TEST         .000       10:20
AIR BLANK            .000       10:20
```

DVM=

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED

INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

284
47

Tucson Police Department Crime Laboratory

364
IJ

SN 66-005003 02/15/06
E767.14 10:22
INVALID TEST

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

ODM
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
02/15/06 10:23

TEST	BRAC VALUE	TIME
AIR BLANK	.000	10:23
INVALID SAMPLE	.MOA	10:23
AIR BLANK	.000	10:23

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

ODM
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

494
I-J

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005003
02/15/06 10:35

TEST	BRAC VALUE	TIME
AIR BLANK	.000	10:34
CAL. CHECK	.044	10:34
INTERFERENT SUBTRACTED		
AIR BLANK	.000	10:35

SUBJECT'S NAME

TIME FIRST OBSERVED INSTRUMENT LOCATION

OPERATOR

SN 66-005003 02/15/06
E767.14 10:32
INVALID TEST

UNABLE TO OBTAIN A
STABLE REFERENCE.

SN 66-005003 02/15/06
E767.14 10:31
INVALID TEST

TEST	BRAC VALUE	TIME
AIR BLANK	.000	10:24
*SUBJECT TEST	.072	10:27

* DEFICIENT SAMPLE - VALUE
PRINTED WAS HIGHEST OBTAINED.

CHECK AMBIENT CONDITIONS

SUBJECT'S NAME

TIME FIRST OBSERVED INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.