EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Rhonda Misse 42071
(Print Name)

INTOXILYZER SERIAL # 66-005004 LOCATION Lab
DATE January 13 19 2004 RM TIME 1747

OK Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

OK DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .414.

OK Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

OK Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

OK Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

OK Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

OK Alcohol free subject Test result 0.000 AC

OK Error recognition logic system functioning
Invalid test printed

OK OK Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

OK OK Calibration standard 0.080 AC
Results: 0.078AC 0.078AC

Instrument operating properly and accurately. YES ✓ NO __

COMMENTS: _____

SIGNATURE Rhonda Misse

0.___ alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005004
01/13/04 17:48

TEST	BRAC VALUE	TIME
AIR BLANK	.000	17:47
SUBJECT TEST	.000	17:47
AIR BLANK	.000	17:48

DVM = .414

SUBJECT'S NAME

Lab

TIME FIRST OBSERVED INSTRUMENT LOCATION

42071

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

SN 66-[illegible] 01/13/04
E767.14 17:58
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

Lab

TIME FIRST OBSERVED INSTRUMENT LOCATION

42071

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

P. 2 of 6
RW

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005004
01/13/04 17:55

TEST	BRAC VALUE	TIME
AIR BLANK	.000	17:54
CAL. CHECK	.078	17:55
AIR BLANK	.000	17:55

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION Lab

42071

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005004
01/13/04 17:54

TEST	BRAC VALUE	TIME
AIR BLANK	.000	17:53
CAL. CHECK	.078	17:53
AIR BLANK	.000	17:54

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION Lab

42071

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

P. 3 of 6

P. 4 of 4
RM

SN [illegible]
E7AP.14
INVALID TEST

[illegible]
18:55

TEST	BRAC VALUE	TIME
AIR BLANK	.000	18:05

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

Lab
INSTRUMENT LOCATION

TIME FIRST OBSERVED

42071
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date Jan 13, 2004 Time 1748
Intoxilyzer Serial# 66-005004 Location Lab
QA Specialist Rhonda Miske 42071
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.078AC, 0.078 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Rhonda Miske

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005004
01/13/04 18:00

TEST	BRAC VALUE	TIME
AIR BLANK	.000	18:00
INVALID SAMPLE	.MOA	18:00
AIR BLANK	.000	18:00

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION Lab

OPERATOR 42071

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005004
01/13/04 18:04

TEST	BRAC VALUE	TIME
AIR BLANK	.000	18:01
*SUBJECT TEST	.072	18:04
AIR BLANK	.000	18:04

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION Lab

OPERATOR 42071

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

P. 4 of 6
RM