



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

*gpc*

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 5000 Serial Number 66-005004 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 9-8-2006 Cal 1 0.079 AC Cal 2 0.079 AC

Date: — Cal 1 — AC Cal 2 — AC

Date: — Cal 1 — AC Cal 2 — AC

Between the dates of N/A and N/A the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 9-21-06.  
The instrument was returned to service on —.


*Terry S. Gallegos*

Terry S. Gallegos

October 30, 2006

ADPS QAS Permit # QAS 976

SUBSCRIBED AND SWORN to before me this 30, day of October, 2006.



*Kathy A. Kinney*

Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

TS4

Date: September 19 2006

QAS: Jennifer Coombs 46160

Intoxilyzer Serial Number: 66-005004

Reason For Service: Dfc. O'Hara told me in person that this instrument was showing a printer error during the initial diagnostics. Terry Gallegos picked up the instrument from the DUI HQ on 9/21/06 and brought it to the lab.

Date Out of Service: 9/21/06

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

Sent to Applied Electronics

---

---

---

---



# MEMORANDUM

---

**DATE:** September 21, 2006

**TO:** Whom it may concern

**FROM:** Jennifer L. Coombs  
Intoxilyzer 5000 QAS  
Tucson Police Department  
Crime Lab  
270 S. Stone Avenue  
Tucson, AZ 85701-1917  
(520) 791-4494

A handwritten signature in black ink, appearing to be "JL Coombs", located to the right of the "FROM:" field.

**SUBJECT:** Intoxilyzer Out of Service

Intoxilyzer 5000, serial number **66-005004** was taken out of service for repair on **September 21, 2006**.

I cannot testify that it was working properly and accurately between the dates of **September 8, 2006** and **September 21, 2006**.

This instrument was located at **DUI HQ**.