EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran
(Print Name)

INTOXILYZER SERIAL #66-005005 LOCATION LAB

DATE January 29 ~~19~~ 2004 TIME 1350

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .512.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
✓ Invalid test printed

✓ Proper sample recognition system
✓ Invalid sample printed
✓ Deficient sample printed

N/A Completeness of sample purge with collection tube**

Calibration standard 0.080 AC

✓ Results: 0.077AC & 0.077AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0. ✓ alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

[illegible]
[illegible] 14:0[illegible]
[illegible]
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

TPO
INT[illegible]
AZ [illegible] SN 66-[illegible]
01/29/04 13:51

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:51
SUBJECT TEST	.000	13:51
AIR BLANK	.000	13:51

DVM = .512

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

p. 286

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
01/29/04 13:55

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:55
CAL. CHECK	.077	13:55
AIR BLANK	.000	13:55

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

LS

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
01/29/04 13:57

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:56
CAL. CHECK	.077	13:56
AIR BLANK	.000	13:57

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

LS

Tucson Police Department Crime Laboratory

p. 386
LS

p 4 of 6

INVALID TEST

SAMPLE INTRODUCED
AT IMPROPER TIME.

DORAN
SUBJECT'S NAME

LAB
TIME FIRST OBSERVED
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

p. 596

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/29/04 Time 150 pm/1350
Intoxilyzer Serial# 66005005 Location LAB
QA Specialist Irene Doran
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.077 AC + 0.077 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

AZ MODEL 5000 SN 66-005005
01/29/04 14:11

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:10
INVALID SAMPLE	.MDA	14:11
AIR BLANK	.000	14:11

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Binaca

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

AZ MODEL 5000 SN 66-005005
01/29/04 14:09

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:06
*SUBJECT TEST	.071	14:09
AIR BLANK	.000	14:09

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

LAB
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P 686