106
PS

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* PM

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)

INTOXILYZER SERIAL # 66-005005 LOCATION DUI HQ
DATE May 19 19 2004 TIME 1333

PS Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

PS DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .545.

PS Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

PS Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

PS* Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

PS Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

PS Alcohol free subject Test result 0.000 AC

PS Error recognition logic system functioning
Invalid test printed PS

PS Proper sample recognition system
Invalid sample printed PS
Deficient sample printed PS

N/A Completeness of sample purge with collection tube**

PS Calibration standard 0.080 AC
Results: 0.081 AC + 0.081 AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS:* Moved ahead 3min 1430-1433
breath hose not heating
changed breath hose, cleaned print roller

SIGNATURE Irene Doran

0.N/A PS alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 66-003005 05/19/04
E767.14 14:49
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |
Q306 |
Q305 |
Q302 |
Q303 |
Q304 |
Q301 |

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 - SN 66-005005
05/19/04 14:37

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:36
SUBJECT TEST	.000	14:37
AIR BLANK	.000	14:37

DVM= .545

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
05/19/04 14:34

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:33
CAL. CHECK	.081	14:34
AIR BLANK	.000	14:34

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P4

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
05/19/04 14:35

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:35
CAL. CHECK	.081	14:35
AIR BLANK	.000	14:35

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P5

Tucson Police Department Crime Laboratory

P4

3
2
6
9

4 of 6

SN 66-005005 05/19/04
E767.14 15:00
INVALID TEST

SAMPLE INTRODUCED
AT IMPROPER TIME.

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

5 8 6
PS

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES TSH
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 5/19/04 Time 1333
Intoxilyzer Serial# 68-005005 Location DUI HQ
QA Specialist Irene Doran (46331)
(Print Name)

(PS) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(PS) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(PS) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(PS) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

PS (PS) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

PS (PS) 6. Push Start Test button or press enter on keyboard.

PS (PS) 7. Insert card in response to display.

PS (PS) 8. Air blank completed.

PS (PS) 9. Calibration check completed.
Test results 0.081 + 0.081 AC

PS (PS) 10. Air blank completed.

PS (PS) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(PS) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
05/19/04 14:47

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:46
INVALID SAMPLE	.MOA	14:46
AIR BLANK	.000	14:47

BEGIN NEW DEPRIVATION PERIOD

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Binaca

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

TPD
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-005005
05/19/04 14:59

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:56
*SUBJECT TEST	.068	14:59
AIR BLANK	.000	14:59

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.