



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

By
11-19-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/6/2007 Cal 1 0.098 AC Cal 2 0.098 AC


Between the dates of 9/26/2007 and 11/6/2007 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah L. Cione
Sarah L. Cione

11/7/2007
ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 7, day of November, 2007.



Kathy A Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
11-19-07

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione 50616 AGENCY Tucson Police Department

DATE 11/6/2007 TIME 13:16

INTOXILYZER SERIAL # 80-000323 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
11/06/2007 13:16:32

Standard Lot#: 706601I
Last Changed By: J. TRAINING #99999

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:18:07
Diagnostic Test	Pass	13:18:34
Air Blank	0.000	13:19:01
0.100 Cal Check	0.099	13:19:20
Air Blank	0.000	13:19:50
Subject Test	0.000	13:20:15
Air Blank	0.000	13:20:43
Five Minute Wait	ABT*	
Air Blank	0.000	13:21:24

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 16-127, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 75C

184

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
11/06/2007 13:22:08

Standard Lot#: 706601I
Last Changed By: J. TRAINING #99999

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:23:11
Air Blank	IPS**	13:23:24

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 15-27, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
11/06/2007

13:23:49

4 of 7 SC
B4

Standard Lot#: 706601I
Last Changed By: J. TRAINING #99999

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:24:40
Diagnostic Test	Pass	13:25:07
Air Blank	0.000	13:25:34
0.100 Cal Check	0.098	13:25:53
Air Blank	0.000	13:26:22
Subject Test	0.024*	13:29:31
Air Blank	0.000	13:30:02
Air Blank	0.000	13:30:30
Subject Test	XXX**	13:30:59
Air Blank	0.000	13:31:32

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 20-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
11/06/2007

13:33:45

5 of 7 sc

Standard Lot#: 706601I
Last Changed By: J. TRAINING #99999

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:34:13
Cal Check	0.098	13:34:31
Air Blank	0.000	13:35:01

Sarah Cione

Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

B4

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione 50616 AGENCY Tucson Police Department
DATE 11/6/07 TIME 13:10
Intoxilyzer Serial # 80-000323 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
11/06/2007

13:10:27

7 of 7 SC

B4

Standard Lot#: 706601I
Last Changed By: J. TRAINING #99999

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:10:55
Cal Check	0.098	13:11:13
Air Blank	0.000	13:11:43

Sarah Cione

Operator's Signature