



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

TS4
8-28-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8/22/08 Cal 1 0.097 AC Cal 2 0.095 AC

Between the dates of 7/25/08 and 8/22/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -
The instrument was returned to service on -


Sarah Cione

Sarah L. Cione

8/27/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 27th, day of August, 20 08.



A.M. Hernandez

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 10 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
8.28.08

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 8/22/08 TIME 13:52

INTOXILYZER SERIAL # 80-000323 LOCATION Crime Lab

- (✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES NO

COMMENTS Had to re-do Calibration Check + first SDAP check because
Standard Lot # was incorrect as a result of a previous in-lab
experiment.
SC

SIGNATURE Sarah Cione

2 of 10 SC

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
08/22/2008 13:54:31

Standard Lot#: SC081208
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:14
Diagnostic Test	Pass	13:55:42
Air Blank	0.000	13:56:09
0.150 Cal Check	0.000*	13:56:51
Air Blank	0.000	13:57:23

*Calibration Check Out of Tolerance

Not a Successfully Completed Test Sequence

3 of 10 sc

104

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
08/22/2008 13:57:43

Standard Lot#: SC081208
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	13:58:14
Air Blank	0.000	13:58:42

*Sequence Aborted

Not a Successfully Completed Test Sequence

4 of 10 sc

154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
08/22/2008 13:59:35

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:00:20
Diagnostic Test	Pass	14:00:47
Air Blank	0.000	14:01:15
0.100 Cal Check	0.097	14:01:33
Air Blank	0.000	14:02:03
Subject Test	0.000	14:02:19
Air Blank	ABT*	14:02:27
Air Blank	0.000	14:02:55

*Sequence Aborted

Not a Successfully Completed Test Sequence

5 of 10 sc

BM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
08/22/2008 14:03:39

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	14:04:31
Air Blank	IPS**	14:04:39

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

6 of 10 sc
BY

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
08/22/2008 14:04:57

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:05:39
Diagnostic Test	Pass	14:06:07
Air Blank	0.000	14:06:34
0.100 Cal Check	0.096	14:06:53
Air Blank	0.000	14:07:22
Subject Test	0.120*	14:10:30
Air Blank	0.000	14:11:03
Air Blank	0.000	14:11:30
Subject Test	XXX**	14:11:54
Air Blank	0.000	14:12:27

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
08/22/2008 14:12:39

7 of 10 SC
104

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	14:13:06
Cal Check	0.095	14:13:25
Air Blank	0.000	14:13:54

Sarah Cione

Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

8 of 10 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

TJH
8-28-08

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 8/22/08 TIME 13:52
Intoxilyzer Serial # 80-000323 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
08/22/2008 13:52:56

9 of 10 SC
134

Standard Lot#: SC081208
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:53:23
Cal Check	0.097	13:53:42
Air Blank	0.000	13:54:11

Sarah Cione
Operator's Signature

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-000323

08/22/2008

14:14:00

10 of 10 SC
BY

Standard Lot#: 630601I

Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616

TUCSON PD

Test	g/210L	Time
Air Blank	0.000	14:14:28
Cal Check	0.096	14:14:46
Air Blank	0.000	14:15:16



Operator's Signature