



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

TS4
11.4.08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10-17-08 Cal 1 0.097 AC Cal 2 0.095 AC

Between the dates of 9-19-08 and 10-17-08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley


11-4-08

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 4th, day of November, 2008.

Am Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 cu

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSG
11.4.08

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department
DATE 10-17-08 TIME 1320
INTOXILYZER SERIAL # 80-000323 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES NO

COMMENTS None.

SIGNATURE Clare Conley

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
10/17/2008

13:22:19

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:23:03
Diagnostic Test	Pass	13:23:31
Air Blank	0.000	13:23:58
0.100 Cal Check	0.097	13:24:17
Air Blank	0.000	13:24:46
Subject Test	0.000	13:25:04
Air Blank	ABT*	13:25:10
Air Blank	0.000	13:25:38

*Sequence Aborted

Not a Successfully Completed Test Sequence

2867

en

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
10/17/2008

13:25:45

367
Cu
134

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:26:38
Air Blank	IPS**	13:26:54

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
10/17/2008

13:27:07

467
cu

154

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP

DOB: 11/11/1111

Sex: M

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:28:04
Diagnostic Test	Pass	13:28:31
Air Blank	0.000	13:28:59
0.100 Cal Check	0.097	13:29:17
Air Blank	0.000	13:29:47
Subject Test	0.132*	13:32:55
Air Blank	0.000	13:33:28
Air Blank	0.000	13:33:55
Subject Test	XXX**	13:34:18
Air Blank	0.000	13:34:52

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
10/17/2008 13:35:10

5067
cu
154

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Test g/210L Time

Air Blank 0.000 13:35:37
Cal Check 0.095 13:35:56
Air Blank 0.000 13:36:26

cu
~~0.095~~ 0.000 *C. Conley*

Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

687
cu
TS4
11-4-08

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 10-17-08 TIME 1320

Intoxilyzer Serial # 80-000 323 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
10/17/2008

13:20:19

Standard Lot#: 630601I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:20:46
Cal Check	0.097	13:21:05
Air Blank	0.000	13:21:34

Clare Conley

Operator's Signature

787 Cu

154