CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TJH 4.13.2010

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 04/08/10 Cal 1 0.099 AC Cal 2 0.097 AC

Between the dates of 03/11/10 and 04/08/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

04/13/10

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 13th, day of April, 2010.

A.M. Hernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

127 CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 04/08/10 TIME 12:22

INTOXILYZER SERIAL # 80-000323 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

227 cs

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
04/08/2010 12:23:51

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:24:40
Diagnostic Test	Pass	12:25:08
Air Blank	0.000	12:25:35
0.100 Cal Check	0.099	12:25:54
Air Blank	0.000	12:26:23
Subject Test	0.000	12:26:43
Air Blank	ABT*	12:26:50
Air Blank	0.000	12:27:18

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 8 7
cs

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
04/08/2010 12:27:22

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:27:56
Air Blank	IPS**	12:28:04

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

427
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
04/08/2010 12:28:08

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:28:56
Diagnostic Test	Pass	12:29:24
Air Blank	0.000	12:29:51
0.100 Cal Check	0.099	12:30:10
Air Blank	0.000	12:30:39
Subject Test	0.135*	12:33:47
Air Blank	0.000	12:34:20
Air Blank	0.000	12:34:48
Subject Test	XXX**	12:35:05
Air Blank	0.000	12:35:40

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

587
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
04/08/2010 12:36:04

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:36:32
Cal Check	0.097	12:36:50
Air Blank	0.000	12:37:20

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 7
CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Cheryl Straub AGENCY Tucson Police Department
DATE 04/08/10 TIME 12:22
Intoxilyzer Serial # 80-000323 Location Crime Lab

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.

OR

Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

737
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
04/08/2010 12:22:29

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:22:56
Cal Check	0.099	12:23:15
Air Blank	0.000	12:23:44

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

182
CS

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

TJH 4-13-2010

Date: 04/08/10

QAS: Cheryl STRAUB

Intoxilyzer Serial Number: 80-000323

New Lot Number: 642845

New Expiration Date: 07/23/2011

(✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 2. Remove the old gas tank and install the new one.

(✓) 3. Hit the "ESC" key twice and swipe the QAS card.

(✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:

Enter Target Value: 0.100
Enter Control Lot #: From the tank label
Enter Exp Date: From the tank label (mm/dd/yyyy)
Edit info Y/N? Y or N
Update Std Info? Y ENTER
"Please wait saving…"

(✓) 5. Hit the "ESC" key twice.

(✓) 6. Go to the "M-Maintenance" menu and select "Control Testing. " Select "D" for dry control test. After selection is made press ENTER.

(✓) 7. Air blank completed.

(✓) 8. Calibration check completed. Test results 0.099 AC.

(✓) 9. Air blank completed.

(✓) 10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 2
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
04/08/2010 12:20:51

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:21:18
Cal Check	0.099	12:21:37
Air Blank	0.000	12:22:06

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.