



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

[Handwritten signature]

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12.1.2006 Cal 1 0.099 AC Cal 2 0.098 AC

The instrument was placed into service on 12.1.2006.


[Handwritten signature: Terry S. Gallegos]

Terry S. Gallegos

1.4.2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 4, day of January, 2007.



[Handwritten signature: Kathy A. Kinney]

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE *je*

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 12-1-2006 TIME 1438
INTOXILYZER SERIAL # 80-000332 LOCATION Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Terry Gallegos

2058
739

INTOXILYZER 8000
Location: TUCSON POLICE DEPT *je*
Serial Number : 80-000332
12/01/2006 14:40:02

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:41:06
Diagnostic Test	Pass	14:41:34
Air Blank	0.000	14:42:01
0.100 Cal Check	0.099	14:42:20
Air Blank	0.000	14:42:50
Subject Test	0.000	14:44:13
Air Blank	0.000	14:44:40
Five Minute Wait	ABT*	
Air Blank	0.000	14:49:26

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 8
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT *je*
Serial Number : 80-000332
12/01/2006 14:49:52

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:50:33
Air Blank	0.000	14:51:01

*Improper Sample

Not a Successfully Completed Test Sequence

4098
754

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
12/01/2006 14:51:54

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:52:48
Diagnostic Test	Pass	14:53:15
Air Blank	RFI*	14:53:26
Air Blank	0.000	14:53:55

*RFI Detect

Not a Successfully Completed Test Sequence

5098 T84

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
12/01/2006 14:55:09

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:56:08
Diagnostic Test	Pass	14:56:35
Air Blank	0.000	14:57:02
0.100 Cal Check	0.099	14:57:21
Air Blank	0.000	14:57:51
Subject Test	0.222*	15:00:59
Air Blank	0.000	15:01:33
Air Blank	0.000	15:02:00
Subject Test	XXX**	15:02:27
Air Blank	0.000	15:03:04

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 8
TSK

INTOXILYZER 8000
Location: TUCSON POLICE DEPT *je*
Serial Number: 80-000332
12/01/2006 15:05:06

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	15:05:33
Cal Check	0.098	15:05:52
Air Blank	0.000	15:06:22


Operator's Signature

70f8
B4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE *je*

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 12-1-2006 TIME 1438
Intoxilyzer Serial # 80-000332 Location Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *Terry Gallegos*

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8078
Tg

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
12/01/2006 14:37:22

Standard Lot#: 502602I
Last Changed By: J. COOMBS #46160

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	14:37:49
Cal Check	0.099	14:38:08
Air Blank	0.000	14:38:38


Operator's Signature



TUCSON POLICE DEPARTMENT
CRIME LABORATORY *je*

INTOXILYZER SERVICE RECORD

1 of 1
TS4

Intoxilyzer Serial Number: 80-000332

Initial Service Date: 12.1.2006

CMI, Inc. Certificate of Calibration Date: 11.7.2003