CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-200332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program. TSY 1-19-2010

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1-7-10 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 12-10-09 and 1-7-10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

1-14-10

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 14, day of January, 20 10.

OFFICIAL SEAL
MONICA L. BRAVO
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. June 27, 2011

Monica L Bravo

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7
CU

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TSY 1.19.2010

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 1-7-10 TIME 0837

INTOXILYZER SERIAL # 80-000332 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
CU

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
01/07/2010 08:42:08

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

BY

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:42:54
Diagnostic Test	Pass	08:43:22
Air Blank	0.000	08:43:49
0.100 Cal Check	0.097	08:44:08
Air Blank	0.000	08:44:37
Subject Test	DEF*	08:47:46
Air Blank	0.000	08:48:13
Air Blank	0.000	08:48:40
Subject Test	0.000	08:49:04
Air Blank	ABT**	08:49:11
Air Blank	RFI***	08:49:25

*Deficient Sample
**Sequence Aborted
***RFI Detect

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7
CM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
01/07/2010 08:49:49

Pb4

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:50:34
Air Blank	IPS**	08:50:43

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

4 of 7
Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
01/07/2010 08:51:22

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:52:15
Diagnostic Test	Pass	08:52:42
Air Blank	0.000	08:53:09
0.100 Cal Check	0.097	08:53:28
Air Blank	0.000	08:53:58
Subject Test	XXX*	08:54:20
Air Blank	0.000	08:54:54

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

5687
CM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
01/07/2010 08:55:16

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:55:43
Cal Check	0.097	08:56:02
Air Blank	0.000	08:56:31

Clare Conley
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6887
CM

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TJH
1.19.2010

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-7-10 TIME 0837

Intoxilyzer Serial # 80.000332 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

787
CN

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
01/07/2010 08:37:41

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

134

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:38:08
Cal Check	0.097	08:38:27
Air Blank	0.000	08:38:57

Clare Conley
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department