



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

mtb

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1.29.2007 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 12.28.2006 and 1.29.2007 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on .
The instrument was returned to service on .

Terry Gallegos
Terry S. Gallegos

2.27.2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 27, day of February, 2007.



Kathy A Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 9
TJG

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

Date: 1.29.2007
 QAS: Terry Gallegos 28847
 Intoxilyzer Serial Number: 80-000.334
 New Lot Number: 627101G
 New Expiration Date: 10.03.2008

- (✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 2. Remove the old gas tank and install the new one.
- (✓) 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change."
 Select "D" for dry gas and enter the following information:
 Enter Target Value: 0.100
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."
- (✓) 5. Hit the "ESC" key twice.
- (✓) 6. Go to the "M-Maintenance" menu and select "T-Control Testing."
 Select "D" for dry control test. After selection is made press
 ENTER.
- (✓) 7. Air blank completed.
- (✓) 8. Calibration check completed. Test results 0. 099 AC.
- (✓) 9. Air blank completed.
- (✓) 10. Remove and sign printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
01/29/2007 17:35:15

2009
TS4

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

~~AZ DPS~~ **TS4** **POLICE DEPARTMENT/CRIME LAB**

Test	g/210L	Time
Air Blank	0.000	17:35:42
Cal Check	0.099	17:36:01
Air Blank	0.000	17:36:31


Operator's Signature

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

700
3 of 9
D4

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 28017 AGENCY Tucson Police Department

DATE 1.29.2007 TIME 1737

INTOXILYZER SERIAL # 80-000334 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓✓✓
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000334
01/29/2007

17:39:55

no
4 of 9
TS4

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

~~AZ DPS~~ TS4 POLICE DEPARTMENT/CRIME LAB

Operator: T. GALLEGOS #28847

~~AZ DPS~~ TS4 POLICE DEPARTMENT/CRIME LAB

Subject: TEST,TEST

DOB: 11/11/1111

Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:41:58
Diagnostic Test	Pass	17:42:25
Air Blank	0.000	17:42:53
0.100 Cal Check	0.100	17:43:11
Air Blank	0.000	17:43:42
Subject Test	0.000	17:44:02
Air Blank	0.000	17:44:29
Five Minute Wait	ABT*	
Air Blank	IPS**	17:44:59

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000334
01/29/2007 17:45:19

5 of 9
TS4

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
~~AZ-DPS~~ TS4 POLICE DEPARTMENT/CRIME LAB

Operator: T. GALLEGOS #28847
~~AZ-DPS~~ TS4 POLICE DEPARTMENT/CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:46:19
Diagnostic Test	Pass	17:46:47
Air Blank	0.000	17:47:14
0.100 Cal Check	INT*	17:47:33
Air Blank	RFI**	17:47:50

*Interferent Detect
**RFI Detect

due to Radio during the
cal check
TS4

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000334
01/29/2007 17:48:18

ms
6 of 9
TS4

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

~~AZ DPS~~ **TS4** POLICE DEPARTMENT/CRIME LAB

Operator: T. GALLEGOS #28847

~~AZ DPS~~ **TS4** POLICE DEPARTMENT/CRIME LAB

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:49:10
Diagnostic Test	Pass	17:49:37
Air Blank	0.000	17:50:05
0.100 Cal Check	0.099	17:50:23
Air Blank	0.000	17:50:54
Subject Test	0.156*	17:54:02
Air Blank	0.000	17:54:35
Air Blank	0.000	17:55:02
Subject Test	XXX**	17:55:27
Air Blank	0.000	17:56:01

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

710
8 of 9
TS4

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 1-29-2007 TIME 1737
Intoxilyzer Serial # 80-000334 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-000334

01/29/2007

17:37:49

Standard Lot#: 627101G

Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

~~AZ DPS~~ **TS4** **POLICE DEPARTMENT/CRIME LAB**

203
9 of 9
TS4

Test	g/210L	Time
Air Blank	0.000	17:38:17
Cal Check	0.100	17:38:36
Air Blank	0.000	17:39:06


Operator's Signature