CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

TSH 8.1.08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000712 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8/1/08 Cal 1 0.099 AC Cal 2 0.097 AC

Between the dates of 7/3/08 and 8/1/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on –.
The instrument was returned to service on –.

Sarah Cione

Sarah L. Cione

8/1/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 1st, day of August, 20 08.

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2011

AM Hernández

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
8.1.08

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 8/1/08 TIME 10:23

INTOXILYZER SERIAL # 80-000712 LOCATION Vehicle # 866

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Sarah Cione

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7 sc

B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
08/01/2008 10:24:57

Standard Lot#: 7354016
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:25:55
Diagnostic Test	Pass	10:26:24
Air Blank	0.000	10:26:51
0.100 Cal Check	0.098	10:27:10
Air Blank	0.000	10:27:39
Subject Test	0.000	10:28:01
Air Blank	0.000	10:28:28
Five Minute Wait	ABT*	
Air Blank	0.000	10:29:05

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1227, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7 SC

TSH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
08/01/2008 10:29:44

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:30:14
Air Blank	IPS**	10:30:22

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 SC

TSH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
08/01/2008 10:30:28

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:31:17
Diagnostic Test	Pass	10:31:46
Air Blank	0.000	10:32:13
0.100 Cal Check	0.098	10:32:32
Air Blank	0.000	10:33:01
Subject Test	0.497*	10:36:10
Air Blank	0.000	10:36:43
Air Blank	0.000	10:37:10
Subject Test	XXX**	10:37:33
Air Blank	0.000	10:38:06

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7 SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
08/01/2008 10:38:28

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:38:56
Cal Check	0.097	10:39:14
Air Blank	0.000	10:39:44

Sarah Cione
Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

6 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

PJ 8.1.08

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 8/1/08 TIME 10:23

Intoxilyzer Serial # 80-000712 Location Vehicle # 866

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7 SC

TBG

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
08/01/2008 10:23:03

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:23:30
Cal Check	0.099	10:23:49
Air Blank	0.000	10:24:18

Sarah Cione
Operator's Signature

Pursuant to section 28-1227, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.