EXHIBIT G-4

1 712 CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSH 7.28.09

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 07/28/09 TIME 12:37

INTOXILYZER SERIAL # 80-000712 LOCATION Crime LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS _____

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 12
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/28/2009 12:40:51

B4

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:41:49
Diagnostic Test	Pass	12:42:16
Air Blank	0.000	12:42:44
0.100 Cal Check	0.099	12:43:02
Air Blank	0.000	12:43:32
Subject Test	0.000	12:44:02
Air Blank	ABT*	12:44:10
Air Blank	0.000	12:44:39

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3812
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/28/2009 12:45:12

TSH

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:45:52
Air Blank	IPS**	12:46:01

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4912 cs

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/28/2009 12:46:28

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:47:21
Diagnostic Test	Pass	12:47:49
Air Blank	0.000	12:48:16
0.100 Cal Check	0.100	12:48:35
Air Blank	0.000	12:49:04
Subject Test	0.177*	12:52:13
Air Blank	0.000	12:52:45
Air Blank	0.000	12:53:12
Subject Test	XXX**	12:53:39
Air Blank	0.000	12:54:13

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 12
cs

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
07/28/2009 12:54:47

T54

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:55:14
Cal Check	0.098	12:55:33
Air Blank	0.000	12:56:03

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 9 12 cs

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSH 7-28-09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department

DATE 07/28/09 TIME 12:37

Intoxilyzer Serial # 80-000712 Location Crime Lab

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7712 cs

B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
07/28/2009 12:37:37

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:38:05
Cal Check	0.099	12:38:24
Air Blank	0.000	12:38:53

Cheryl Stu

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 of 12 CS

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

TSH 7-28-09

INTOXILYZER SERVICE RECORD

QAS: Cheryl Straub

Intoxilyzer Serial Number: 80-000712

Date Returned to Service: 07/28/09

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

9 9 12 c8

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD FOR CALIBRATION GAS TANK CHANGE

TSH 7.28.09

Date: 07/28/09

QAS: Cheryl STRAUB

Intoxilyzer Serial Number: 80-000712

New Lot Number: 82050II

New Expiration Date: 29 July 2010

(✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 2. Remove the old gas tank and install the new one.

(✓) 3. Hit the "ESC" key twice and swipe the QAS card.

(✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:

Enter Target Value: 0.100
Enter Control Lot #: From the tank label
Enter Exp Date: From the tank label (mm/dd/yyyy)
Edit info Y/N? Y or N
Update Std Info? Y ENTER
"Please wait saving..."

(✓) 5. Hit the "ESC" key twice.

(✓) 6. Go to the "M-Maintenance" menu and select "Control Testing. " Select "D" for dry control test. After selection is made press ENTER.

(✓) 7. Air blank completed.

(✓) 8. Calibration check completed. Test results 0.099 AC.

(✓) 9. Air blank completed.

(✓) 10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

10 of 12 CS

TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
07/28/2009 11:35:03

Standard Lot#: 820501I
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:35:31
Cal Check	0.099	11:35:50
Air Blank	0.000	11:36:19

Cheryl Straub
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

11912 cs

CMI INC.

316 E. 9th St., Owensboro, KY 42303
270-685-6200 — Fax 270-685-6288
1-800-835-0690

SERVICE WORK ORDER

32 0215 BH

INTOX
PBT
TOXITEST

DATE: [illegible]
RECEIVED VIA: [illegible]

BILL TO

ATTN:
[illegible]
CITY: [illegible] STATE: [illegible] ZIP: [illegible]
CUSTOMER NO.
PHONE

SHIP TO

[illegible]
ATTN: [illegible]
[illegible]
CITY: [illegible] STATE: [illegible] ZIP: [illegible]
PHONE

MODEL #	S/N	DATE RCVD.	P.O. #	REQ.	BILL CODE	SERVICE LOCATION CODE
[illegible]	[illegible]					

PROD. CODE
EXTRA PARTS RCVD.: [illegible]

DESCRIPTION OF PROBLEM: PRINTER JAM + OPEN BREATH TUBE.

☐ QUOTE
WARRANTY ☐ SERVICE DEPT. ☐ NEW

☑ PRELIMINARY TEST ☑ DISMANTLE, INSPECT, DIAGNOSTIC TEST
POWER SUPPLY ☑ TEST, INSPECT ☐ REPLACE
☑ MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
PWB's ☐ DIAGNOSE, REPAIR ☐ REPLACE
☐ PWB's - SURFACE CLEAN IN PLACE
☐ FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
CHOPPER WHEEL MOTOR ☐ TEST & INSPECT ☐ REPLACE
☑ CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
IR LIGHT SOURCE ☐ REMOVE, REALIGN ☑ REPLACE
PRINTER ASSY. ☐ DISASSEMBLE ☐ CLEAN & LUBRICATE
☐ REMOVE & REPLACE BELT ☐ REALIGNMENT
☐ REPLACE ☐ TEST PRINTER FUNCTIONS
BATTERIES - REPLACE ☐ CPU ☐ RAM
☑ REPLACE PARTS ☑ REASSEMBLE ☑ CALIBRATION ☑ FINAL TEST
☐ MISC.
PBT ☐ PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
☐ DISASSEMBLE ☐ REPLACE PARTS ☐ REASSEMBLE
☐ TEST ☐ CALIBRATE ☐ FINAL TEST
TOXITEST ☐ INCOMING LEAK TEST ☐ VISUAL/MECH. INSPECTION
☐ TEST THERMOMETER & THERMOSTAT ☐ DISASSEMBLE
☐ REPLACE PARTS ☐ REASSEMBLE ☐ OUTGOING & LEAK TEST
☐ FINAL TEST
☐ MISC.

PARTS USED

QTY	ID NO.	DESCRIPTION	CODE
1	021307	Breath tube	
1	021297	Pump	
1	020982	Source	
1	210120	Flow lenses	

COMMENTS: Replaced listed parts Checked out.

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

SEAL # 02191

HOURS: 4.0

TECH. [illegible] DATE [illegible] CAL. TECH. [illegible] DATE [illegible]

FORM # CMI 8WO 1 REV (-)

SHIPPING DATE/CHARGES

SPECIAL SHIPPING INSTRUCTIONS

☐ COD ☐ AMOUNT $

12 of 12
CS

CERTIFICATE OF CALIBRATION

INTOXILYZER® *8000*

This is to certify that the calibration of INTOXILYZER serial number 80-000712, *manufactured by CMI, subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST (NBS) standard material 1828.*

Date 7/7/09 *Signed* Kristi Robinson
Technician

CMI INC.

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

316 EAST 9TH STREET
OWENSBORO, KY. 42303