



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
7-25-2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000712 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7-21-11 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 6-23-11 and 7-21-11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley


7-21-11

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 21st day of July, 2011.

Susan A. Hitchye

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

1067
cu

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TS4
7-25-2011

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department
DATE 7-21-11 TIME 0844
INTOXILYZER SERIAL # 80-000712 LOCATION ODW

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS None.

SIGNATURE Clare Conley

EXHIBIT G-2

207
m

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

BSG
7-25-2011

QA Specialist Clare Conley AGENCY Tucson Police Department
DATE 7-21-11 TIME 0844
Intoxilyzer Serial # 80-000712 Location DDW

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
07/21/2011 08:44:00


3507
CM

B4

Standard Lot#: 719919
Last Changed By: S. CLEARY #44524

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:44:28
Cal Check	0.097	08:44:47
Air Blank	0.000	08:45:16



Operator's Signature

4567
CW

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/21/2011 08:45:43

Standard Lot#: 719919
Last Changed By: S. CLEARY #44524

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:46:27
Diagnostic Test	Pass	08:46:55
Air Blank	0.000	08:47:22
0.100 Cal Check	0.097	08:47:41
Air Blank	0.000	08:48:10
Subject Test	0.000	08:48:38
Air Blank	ABT*	08:48:44
Air Blank	0.000	08:49:12

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5067
CR
134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/21/2011 08:49:17

Standard Lot#: 719919
Last Changed By: S. CLEARY #44524

OAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:50:10
Diagnostic Test	Pass	08:50:37
Air Blank	0.000	08:51:04
0.100 Cal Check	0.100	08:51:23
Air Blank	0.000	08:51:53
Subject Test	RFI*	08:52:06
Air Blank	IPS**	08:52:15

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000712
07/21/2011 08:52:46

6-67
CW

BSH

Standard Lot#: 719919
Last Changed By: S. CLEARY #44524

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:53:38
Diagnostic Test	Pass	08:54:05
Air Blank	0.000	08:54:32
0.100 Cal Check	0.097	08:54:51
Air Blank	0.000	08:55:21
Subject Test	0.095*	08:58:29
Air Blank	0.000	08:59:00
Air Blank	0.000	08:59:28
Subject Test	XXX**	08:59:48
Air Blank	0.000	09:00:21

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INFOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-000712

07/21/2011

09:01:16

7567
m

Standard Lot#: 719919

Last Changed By: S. CLEARY #44524

154

OAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:01:44
Cal Check	0.097	09:02:02
Air Blank	0.000	09:02:32

Clare Conley

Operator's Signature