CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

TBG 1-11-2008

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1-10-08 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 12-14-07 and 1-10-08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

1-10-08

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 10th, day of January, 2008.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

1 of 7 Cu

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

DM 1.11.2008

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 1-10-08 TIME 08:02

INTOXILYZER SERIAL # 50-001108 LOCATION ODM

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.098 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7

Au

TBH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001108
01/10/2008 08:03:31

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:04:24
Diagnostic Test	Pass	08:04:51
Air Blank	0.000	08:05:19
0.100 Cal Check	0.100	08:05:37
Air Blank	0.000	08:06:07
Subject Test	0.000	08:06:25
Air Blank	ABT*	08:06:33
Air Blank	0.000	08:07:01

*Sequence Aborted

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Not a Successfully Completed Test Sequence

3 of 7
Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001108
01/10/2008 08:07:06

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:07:38
Air Blank	IPS**	08:07:46

*RFI Detect
**Improper Sample

Pursuant to section 28-1327, Arizona Revised Statutes, this document is [illegible] of the information contained in computer storage devices of the Tucson Police Department.

Not a Successfully Completed Test Sequence

4 of 7
Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001108
01/10/2008 08:08:05

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:08:56
Diagnostic Test	Pass	08:09:23
Air Blank	0.000	08:09:51
0.100 Cal Check	0.100	08:10:09
Air Blank	0.000	08:10:39
Subject Test	0.098*	08:13:47
Air Blank	0.000	08:14:20
Air Blank	0.000	08:14:47
Subject Test	XXX**	08:15:17
Air Blank	0.000	08:15:52

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a [illegible] of the information contained in computer storage devices of the Tucson Police Department.

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001108
01/10/2008 08:16:16

5 of 7 Ch

Standard Lot#: 7198011
Last Changed By: C. CONLEY #34517

TS4

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:16:44
Cal Check	0.098	08:17:02
Air Blank	0.000	08:17:32

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-10-08 TIME 08:02

Intoxilyzer Serial # 80-001108 Location ODM

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001108
01/10/2008 08:02:11

7 of 7
CM

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

TB4

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:02:38
Cal Check	0.099	08:02:57
Air Blank	0.000	08:03:27

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.