



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

DSY
3-11-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 03/11/10 Cal 1 0.103 AC Cal 2 0.101 AC

Between the dates of 02/11/10 and 03/11/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub


03/11/10

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 11th, day of March, 2010.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

177
CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

By
3-11-2010

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department
DATE 03/11/10 TIME 06:06
INTOXILYZER SERIAL # 80-001108 LOCATION TEAM 4

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 101 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Cheryl Straub

287
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001108
03/11/2010

06:09:16

BY

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 06:10:09 |
| Diagnostic Test | Pass | 06:10:37 |
| Air Blank | 0.000 | 06:11:04 |
| 0.100 Cal Check | 0.102 | 06:11:23 |
| Air Blank | 0.000 | 06:11:52 |
| Subject Test | 0.000 | 06:12:19 |
| Air Blank | ABT* | 06:12:26 |
| Air Blank | 0.000 | 06:12:55 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001108
03/11/2010 06:13:08

TS4

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 06:13:52 |
| Air Blank | IPS** | 06:14:04 |

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

477
es

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001108
03/11/2010

06:14:18

BM

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 06:15:12 |
| Diagnostic Test | Pass | 06:15:39 |
| Air Blank | 0.000 | 06:16:07 |
| 0.100 Cal Check | 0.103 | 06:16:25 |
| Air Blank | 0.000 | 06:16:55 |
| Subject Test | 0.102* | 06:20:03 |
| Air Blank | 0.000 | 06:20:35 |
| Air Blank | 0.000 | 06:21:02 |
| Subject Test | XXX** | 06:21:33 |
| Air Blank | 0.000 | 06:22:08 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

577
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-001108
03/11/2010

06:22:29

754

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 06:22:57 |
| Cal Check | 0.101 | 06:23:15 |
| Air Blank | 0.000 | 06:23:45 |



Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

By
3-11-2010

QA Specialist CHERYL STRAUB AGENCY Tucson Police Department
DATE 03/11/10 TIME 06:06
Intoxilyzer Serial # 80-001108 Location TEAM 4

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

777
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-001108
03/11/2010

06:06:22

154

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 06:06:49 |
| Cal Check | 0.103 | 06:07:08 |
| Air Blank | 0.000 | 06:07:38 |



Operator's Signature