



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

TSY  
10-26-2009

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10/15/09 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 9/17/09 and 10/15/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.  
The instrument was returned to service on —.

Sarah Cione

Sarah L. Cione

10/21/09

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 21<sup>st</sup>, day of October, 2009.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

TSy  
10-26-2009

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 10/15/09 TIME 13:50

INTOXILYZER SERIAL # 80-001163 LOCATION Trm 1-Vehicle #944

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS SQAP performed in vehicle with E85 fuel.  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Sarah Cione

2 of 7sc

TS4

INTOXILYZER 800V  
Location: ARIZONA  
Serial Number: 807011163  
10/15/2009 13:51:36

Standard Lot#: 8260011  
Last Changed By: C. STRAUB #50772

Officer: S. CIONE #50616  
TUJSON PD

Operator: S. CIONE #50616  
TUJSON PD

Subject: S2AP, TEST  
DOB: 10/11/1981  
Sex: F Weight: 111

15 Minute Dephivation Period? Yes

Test	g/dl	Time
Air Blank	0.000	13:52:24
Diagnostic Test	Pass	13:52:53
Air Blank	0.000	13:53:20
0.100 Cal Check	0.099	13:53:39
Air Blank	0.000	13:54:09
Subject Test	0.000	13:54:24
Air Blank	0.000	13:54:51
Five Minute Wait	ABT*	
Air Blank	0.000	13:55:26

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

TS4

INTOXICIZER 8000  
 Location: 4P/2044  
 Serial Number: 60-001163  
 10/15/2005 13:55:32

---

Standard Lot#: 8260011  
 Last Changed By: C. STRAUB #50772

---

QAS: S. CIONE #50616  
 TUCSON PD

---

Operator: S. CIONE #50616  
 TUCSON PD

---

Subject: 6049 TEST  
 DOB: 11/11/1111  
 Sex: F Weight: 111

---

15 Minute Depriuation Period? Yes

---

Test:	g/21.0L	Time
Alc Blank	RF1+	13:56:02
RF1 Blank	RF5++	13:56:10

\*RF1 Detect  
 \*\*Improper Sample

Not a Successfully  
 Completed  
 Test Sequence

4 of 7 SC

TS4

INTENT: 0000  
Location: ARIZONA  
Serial Number: 67-00103  
12/15/2015 11:56:15

---

Standard Lot: 626201  
Last Changed By: STABLE #50772

---

CAS: S. CIONE #50615  
TUCSON PD

---

Operator: S. CIONE #50615  
TUCSON PD

---

Subject: SOAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:57:00
Diagnostic Test	Pass	13:57:29
Air Blank	0.000	13:57:56
2.100 Cal Check	2.100	13:58:15
Air Blank	0.000	13:58:45
Subject Test	0.023*	14:01:53
Air Blank	0.000	14:02:25
Air Blank	0.000	14:02:53
Subject Test	XXXXX	14:03:17
Air Blank	0.000	14:03:41

---

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

5 of 7<sub>sc</sub>

154

INTONILY222 8000  
Location: ARIZONA  
Serial Number: 62-001163  
10/15/2013 14:14:00

---

Standard ID#: 8250111  
Last Changed By: C. STRAUB #50772

---

OP#: S. CIONE #50616  
TUCCON PD

---

Test	gr213L	Time
Hit Blank	1.111	14:14:31
Call Check	1.099	14:14:50
Hit Blank	0.000	14:15:00

---

*Sarah Cione*  
Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

6 of 75c

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TS4  
10-26-09

QA Specialist Sarah Cione AGENCY Tucson Police Department  
DATE 10/15/09 TIME 13:50  
Intoxilyzer Serial # 80-001163 Location Tm 1 - Vehicle # 944

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- ( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

7 of 7 sc

154

INDICATED 0000  
Location: 491204  
Serial Number: 80-101163  
10/15/2005 13:50:13

---

Standard Lot#: 8260011  
Last Changed By: C. STRAUB #51772

---

OPERATOR: CIONE #50615  
TUESDAY

---

Test	g 2.0	Time
Air Blank	0.000	13:51:45
Cal. Check	0.100	13:51:59
Air Blank	0.000	13:51:29

---

*Sarah Cione*  
Operator's Signature