



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

TK
6-19-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001165 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6-7-07 Cal 1 0.098 AC Cal 2 0.099 AC

Between the dates of 5-10-07 and 6-7-2007 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.


Clare Conley

Clare L. Conley

June 19, 2007

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 19, day of June, 2007.



Kathy A Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

904 cu
1 of 10 cu

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TSY
6-18-07

QA SPECIALIST Clare Conley AGENCY Tucson Police Department
DATE 6-7-07 TIME 0939
INTOXILYZER SERIAL # 80-001165 LOCATION ODS

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 09100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS Deficient sample check was not performed as part
of the test sequence. When this was discovered, it was
performed on 6-14-07.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 9 cu cu
2 of 10 cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/07/2007 09:41:42

B4
6-18-07

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:42:48
Diagnostic Test	Pass	09:43:15
Air Blank	0.000	09:43:42
0.100 Cal Check	0.098	09:44:01
Air Blank	0.000	09:44:31
Subject Test	0.000	09:44:52
Air Blank	0.000	09:45:19
Five Minute Wait	ABT*	
Air Blank	0.000	09:45:59

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

~~3 of 9 Cu~~
3 of 10 Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/07/2007 09:46:24

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

By
6-18-07

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:47:24
Diagnostic Test	Pass	09:47:51
Air Blank	0.000	09:48:18
0.100 Cal Check	0.098	09:48:37
Air Blank	IPS*	09:48:44
Air Blank	0.000	09:49:12

*Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

of 9 cu
cu

4 of 10 cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/07/2007 09:49:39

T84
6-18-07

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:50:34
Diagnostic Test	Pass	09:51:01
Air Blank	0.000	09:51:29
0.100 Cal Check	0.098	09:51:47
Air Blank	RFI*	09:51:54
Air Blank	0.000	09:52:22

*RFI Detect

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

~~5 of 9 cu~~
5 of 10 cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/07/2007 09:53:10

Tsh
6-18-07

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:54:16
Diagnostic Test	Pass	09:54:43
Air Blank	0.000	09:55:11
0.100 Cal Check	0.098	09:55:29
Air Blank	0.000	09:56:00
Subject Test	0.021	09:56:19
Air Blank	0.000	09:56:51
Five Minute Wait		
Air Blank	0.000	10:01:47
Subject Test	XXX*	10:02:23
Air Blank	0.000	10:02:57

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

~~6 of 9 Cu~~
Cu
6 of 10 Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/07/2007 10:03:15

TSy
6-18-07

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:04:21
Diagnostic Test	Pass	10:04:49
Air Blank	0.000	10:05:16
0.100 Cal Check	0.099	10:05:35
Air Blank	0.000	10:06:05
Subject Test	ABT*	10:08:35
Air Blank	0.000	10:09:08

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

~~7 of 9 Cu~~
cu
7 of 10 Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001165
06/07/2007 10:09:42

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

TS4
6-1807

Test	g/210L	Time
Air Blank	0.000	10:10:09
Cal Check	0.098	10:10:28
Air Blank	0.000	10:10:58

Clare Conley

Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

8 of 9 cu
cu
8 of 10 cu

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TS4
6.18.07

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department
DATE 6-7-07 TIME 0939
Intoxilyzer Serial # 80-001165 Location ODS

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 150 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

9 of 9 Cu
Cu
9 of 10 Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001165
06/07/2007 09:39:09

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

TS4
6.18.07

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:39:37
Cal Check	0.099	09:39:55
Air Blank	0.000	09:40:25

Clare Conley

Operator's Signature

10 of 10 cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001165
06/14/2007

06:25:35

TS4
6-18-07

Standard Lot#: 630601I
Last Changed By: J. COOMBS #46160

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	06:26:31
Diagnostic Test	Pass	06:26:58
Air Blank	0.000	06:27:26
0.100 Cal Check	0.099	06:27:44
Air Blank	0.000	06:28:15
Subject Test	0.000*	06:31:23
Air Blank	0.000	06:31:50
Air Blank	0.000	06:32:18
Subject Test	ABT**	06:32:28
Air Blank	0.000	06:32:57

*Deficient Sample
**Sequence Aborted

Not a Successfully Completed Test Sequence