



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

TSY  
7-22-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7-22-10 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 6-24-10 and 7-22-10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

7-22-10


ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 22, day of

July, 2010.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1567  
cu  
By  
7-22-2010

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 7-22-10 TIME 0903

INTOXILYZER SERIAL # 80-001199 LOCATION ODW DUI HQ

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS None.  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Clare Conley

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001199  
07/22/2010

09:05:24

287  
lu  
154

---

Standard Lot#: 642845  
Last Changed By: J. OLIVARES #36552

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	09:06:23
Diagnostic Test	Pass	09:06:51
Air Blank	0.000	09:07:18
0.100 Cal Check	0.099	09:07:37
Air Blank	0.000	09:08:06
Subject Test	0.000	09:08:25
Air Blank	ABT*	09:08:31
Air Blank	0.000	09:09:00

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001199  
07/22/2010

09:10:42

3.87  
cu  
B4

---

Standard Lot#: 642845  
Last Changed By: J. OLIVARES #36552

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	09:11:20
Air Blank	IPS**	09:11:29

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001199  
07/22/2010

09:12:13

Standard Lot#: 642845  
Last Changed By: J. OLIVARES #36552

QAS: C. CONLEY #34517  
TUCSON PD

Operator: C. CONLEY #34517  
TUCSON PD

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:13:02
Diagnostic Test	Pass	09:13:29
Air Blank	0.000	09:13:57
0.100 Cal Check	0.099	09:14:15
Air Blank	0.000	09:14:45
Subject Test	0.083*	09:17:53
Air Blank	0.000	09:18:24
Air Blank	0.000	09:18:52
Subject Test	XXX**	09:19:24
Air Blank	0.000	09:19:57

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

4587  
CW  
TS4

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-001199

07/22/2010

09:20:25

-----  
Standard Lot#: 642845

Last Changed By: J. OLIVARES #36552

-----  
QAS: C. CONLEY #34517

TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	09:20:52
Cal Check	0.099	09:21:11
Air Blank	0.000	09:21:40

*Claire Conley*  
-----  
Operator's Signature

5067  
EW  
B4

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

6587  
cu  
TS4  
7-22-2010

QA Specialist Clare Conley AGENCY Tucson Police Department  
DATE 7-22-10 TIME 0903  
Intoxilyzer Serial # 80-001199 Location ODW DUI HQ

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 099 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-001199

07/22/2010

09:03:30

7587  
en

-----  
Standard Lot#: 642845


Last Changed By: J. OLIVARES #36552

154

-----  
QAS: C. CONLEY #34517

TUCSON PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      09:03:57  
Cal Check                      0.099                      09:04:16  
Air Blank                      0.000                      09:04:45

  
-----  
Operator's Signature