CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSH 11.15.2010

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11-11-10 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 10-14-10 and 11-11-10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

11-11-10

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 11, day of November, 2010.

A.M. Hernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
CM
TSG
11.15.2010

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 11-11-10 TIME 0850

INTOXILYZER SERIAL # 80-001199 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
cu
TDK

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
11/11/2010 08:52:05

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:52:53
Diagnostic Test	Pass	08:53:20
Air Blank	0.000	08:53:47
0.100 Cal Check	0.102	08:54:06
Air Blank	0.000	08:54:36
Subject Test	0.000	08:54:58
Air Blank	ABT*	08:55:05
Air Blank	0.000	08:55:33

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

387
CM
BM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
11/11/2010 08:56:34

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:57:02
Air Blank	IPS**	08:57:11

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7
cu
TSH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
11/11/2010 08:57:33

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:58:20
Diagnostic Test	Pass	08:58:47
Air Blank	0.000	08:59:14
0.100 Cal Check	0.101	08:59:33
Air Blank	0.000	09:00:02
Subject Test	0.227*	09:03:11
Air Blank	0.000	09:03:43
Air Blank	0.000	09:04:10
Subject Test	XXX**	09:04:45
Air Blank	0.000	09:05:18

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7
Cu

TSY

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
11/11/2010 09:05:54

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:06:21
Cal Check	0.100	09:06:40
Air Blank	0.000	09:07:09

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

687
Cu
TSY
11.15.2010

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 11-11-10 TIME 0850

Intoxilyzer Serial # 80-001199 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
11/11/2010 08:50:37

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:51:04
Cal Check	0.100	08:51:23
Air Blank	0.000	08:51:52

7 of 7
LM
TSH

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1 of 2
eh

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

TJH
11.15.2010

Date: OCTOBER 17 2010

QAS: MORAND 24219

Intoxilyzer Serial Number: 80 · 001199

New Lot Number: 719919

New Expiration Date: 02 02 2012

(X) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 2. Remove the old gas tank and install the new one.

(X) 3. Hit the "ESC" key twice and swipe the QAS card.

(X) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:

Enter Target Value: 0.100
Enter Control Lot #: From the tank label
Enter Exp Date: From the tank label (mm/dd/yyyy)
Edit info Y/N? Y or N
Update Std Info? Y ENTER
"Please wait saving..."

(X) 5. Hit the "ESC" key twice.

(X) 6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.

(X) 7. Air blank completed.

( ) 8. Calibration check completed. Test results 0.099 AC.

(X) 9. Air blank completed.

( ) 10. Remove and sign printed record. Attach the record to the completed checklist.

SIGNATURE C. Morand

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 2
CM

84

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
10/17/2010 19:16:46

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. MORAND #24219
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	19:17:14
Cal Check	0.099	19:17:32
Air Blank	0.000	19:18:02

C. Morand
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.