CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSH 2-7-2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 2-3-11 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 1-6-11 and 2-3-11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

2-4-11

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 4th, day of February, 20 11.

OFFICIAL SEAL
MONICA L. BRAVO
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. June 27, 2011

Monica L Bravo

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1&7 CM

154 2.7.2011

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 2-3-11 TIME 0840

INTOXILYZER SERIAL # 80-001199 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
CR

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
02/03/2011 08:42:51

154

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:43:37
Diagnostic Test	Pass	08:44:04
Air Blank	0.000	08:44:32
0.100 Cal Check	0.100	08:44:50
Air Blank	0.000	08:45:20
Subject Test	0.000	08:46:22
Air Blank	ABT*	08:46:28
Air Blank	0.000	08:46:56

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
02/03/2011 08:47:10

3.07
Cu

154

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:47:41
Air Blank	IPS**	08:47:49

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
02/03/2011 08:49:00

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:49:46
Diagnostic Test	Pass	08:50:13
Air Blank	0.000	08:50:41
0.100 Cal Check	0.100	08:50:59
Air Blank	0.000	08:51:29
Subject Test	0.090*	08:54:37
Air Blank	0.000	08:55:09
Air Blank	0.000	08:55:36
Subject Test	XXX**	08:56:15
Air Blank	0.000	08:56:48

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5567
Cm

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
02/03/2011 08:57:14

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:57:41
Cal Check	0.100	08:58:00
Air Blank	0.000	08:58:29

Clare Conley

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7
CM

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

154
2-7-2011

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 2-3-11 TIME 0840

Intoxilyzer Serial # 80-001199 Location DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7867
CM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
02/03/2011 08:40:54

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:41:21
Cal Check	0.100	08:41:40
Air Blank	0.000	08:42:10

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.