



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
3.17.08

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 3/13/08 Cal 1 0.097 AC Cal 2 0.096 AC

Between the dates of 2/14/08 and 3/13/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.


Sarah Cione

Sarah L. Cione

3/14/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 14, day of March, 2008.



Kathy A. Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1 of 75c

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TS4
3/17-08

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 3/13/08 TIME 11:27

INTOXILYZER SERIAL # 80-002201 LOCATION Vehicle # 817

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

2 of 7sc

TS4

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
03/13/2008 11:28:54

Standard Lot#: 7198011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP_TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:29:57
Diagnostic Test	Pass	11:30:26
Air Blank	0.000	11:30:53
0.100 Cal Check	0.097	11:31:12
Air Blank	0.000	11:31:42
Subject Test	0.000	11:31:59
Air Blank	0.000	11:32:26
Five Minute Wait	ABT*	
Air Blank	0.000	11:33:02

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7 sc
B4

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
03/13/2008 11:33:20

Standard Lot#: 7198011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank 0.000 11:34:14
Diagnostic Test Pass 11:34:43
Air Blank RFI* 11:34:51
Air Blank IPS** 11:34:59

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7sc

TS4

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
03/13/2008 11:35:12

Standard Lot#: 7198011
Last Changed By: S. CIONE #50616

DAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:36:07
Diagnostic Test	Pass	11:36:36
Air Blank	0.000	11:37:03
0.100 Cal Check	0.096	11:37:22
Air Blank	0.000	11:37:51
Subject Test	0.111*	11:41:00
Air Blank	0.000	11:41:33
Air Blank	0.000	11:42:01
Subject Test	XXX**	11:42:21
Air Blank	0.000	11:42:55

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 75c

104

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
03/13/2008 11:43:44

Standard Lot#: 7198011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:44:12
Cal Check	0.096	11:44:31
Air Blank	0.000	11:45:01

Sarah Cione
Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7sc

B4

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 3/13/08 TIME 11:27
Intoxilyzer Serial # 80-002201 Location vehicle # 817

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

7 of 7sc

B4

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
03/13/2008 11:27:22

Standard Lot#: 7198011
Last Changed By: S. CIONE #50616

DAS: S/ CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:27:50
Cal Check	0.097	11:28:08
Air Blank	0.000	11:28:38

Sarah Cione
Operator's Signature
