



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

By 12-22-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12/18/08 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 11/20/08 and 12/18/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -
The instrument was returned to service on -

Sarah Cione

Sarah L. Cione


12/22/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 22 day of December, 2008.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

1 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

T34
12-22-08

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department
DATE 12/18/08 TIME 9:30
INTOXILYZER SERIAL # 80-002201 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 75c

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
12/18/2008

09:31:52

Standard Lot#: 730901I
Last Changed By: J. OLIVARES #36552

AM

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:32:40
Diagnostic Test	Pass	09:33:08
Air Blank	0.000	09:33:35
0.100 Cal Check	0.097	09:33:54
Air Blank	0.000	09:34:24
Subject Test	0.000	09:34:41
Air Blank	0.000	09:35:08
Five Minute Wait	ABT*	
Air Blank	0.000	09:35:43

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7 sc

fsy

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
12/18/2008

09:36:11

Standard Lot#: 730901I
Last Changed By: J. OLIVARES #36552

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:36:50
Air Blank	IPS**	09:36:59

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 sc

134

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
12/18/2008 09:37:05

Standard Lot#: 730901I
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:38:09
Diagnostic Test	Pass	09:38:36
Air Blank	0.000	09:39:04
0.100 Cal Check	0.097	09:39:22
Air Blank	0.000	09:39:52
Subject Test	0.072*	09:43:01
Air Blank	0.000	09:43:33
Air Blank	0.000	09:44:00
Subject Test	XXX**	09:44:36
Air Blank	0.000	09:45:10

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7 sc

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
12/18/2008 09:45:30

154

Standard Lot#: 730901I
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:45:57
Cal Check	0.097	09:46:16
Air Blank	0.000	09:46:46

Sarah Cione
Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TS4
12-22-08

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 12/18/08 TIME 9:30
Intoxilyzer Serial # 80-002201 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

7 of 7 sc

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
12/18/2008 09:30:06

Standard Lot#: 730901I
Last Changed By: J. OLIVARES #36552

134

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:30:33
Cal Check	0.097	09:30:52
Air Blank	0.000	09:31:22

Sarah Cione
Operator's Signature