



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
8.23.2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/2/10 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 7/1/10.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


8/23/10

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 23, day of August, 2010.

A.M. Hernández

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2 sc

INTOXILYZER SERVICE RECORD

TS4
8-23-2010

Date: 7/2/10

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-002201

Reason For Service: Regulator base detached from instrument.
Instrument still operating.

Date Out of Service: 7/1/10

Calibration Check completed prior to removal from service

Date: 7/1/10

SQAP completed prior to removal from service

Date: 7/1/10

Corrective Action:

Sent to CMI, Inc.



2 of 2 sc

Service Evaluation Form

This form **MUST** be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

174
8-23-2010

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)
 Name Terry Gallegos Phone: (520) 837-7677
 Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Ship to Address:
Finance / Accounts Payable Tucson Police Dept. Crime Lab
P.O. Box 27450 270 S. Stone Ave.
Tucson, AZ 85726-7450 Tucson, AZ 85701-1917

3. Serial Number: 80-002201 Instrument Model: Intox 8000

4. Detailed Description of Problem:
Regulator base is detached from instrument. Attempted to re-attach w/ new screws, but won't stay secure.

Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
 Purchase Order Number (attach a copy of P.O. if applicable) 20669

Authorized By:
Terry Gallegos Toxicology Coordinator
 Name (Please Print) Title
Terry Gallegos 7-19-2010
 Signature Date

Ship to:
 CMI, Inc.
 Attn: Service Dept.
 316 East Ninth Street
 Owensboro, KY 42303

No, please send estimate before repairs are made.
 Note: An estimate will be faxed before performing any repairs and may cause delays in service.
 An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.