



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

CM
7-10-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/3/08 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 6/6/08 and 7/3/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -
The instrument was returned to service on -

Sarah Cione

Sarah L. Cione


7/3/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 3, day of July, 2008.

Kathy A Kinney

Notary Public



These documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 SC
Ch
7-10-08

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 7/3/08 TIME 10:16

INTOXILYZER SERIAL # 80-002261 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

2 of 75c

Ch

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
07/03/2008 10:18:03

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:18:46
Diagnostic Test	Pass	10:19:14
Air Blank	0.000	10:19:41
0.100 Cal Check	0.101	10:20:00
Air Blank	0.000	10:20:30
Subject Test	0.000	10:20:54
Air Blank	0.000	10:21:21
Five Minute Wait	ABT*	
Air Blank	0.000	10:21:55

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7sc

OK

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
07/03/2008 10:23:18

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:24:03
Air Blank	IPS**	10:24:12

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

4 of 75c

ck

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
07/03/2008 10:27:09

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:28:15
Diagnostic Test	Pass	10:28:42
Air Blank	0.000	10:29:09
0.100 Cal Check	0.101	10:29:28
Air Blank	0.000	10:29:58
Subject Test	0.027*	10:33:06
Air Blank	0.000	10:33:38
Air Blank	0.000	10:34:06
Subject Test	XXX**	10:34:38
Air Blank	0.000	10:35:11

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence


INTOXILYZER 8000
Location: CMI INC
Serial Number: 80-002261
07/03/2008

10:35:48

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:36:16
Cal Check	0.100	10:36:35
Air Blank	0.000	10:37:04



Operator's Signature

5 of 75c

OK

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104 (A)

6 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

CM 7-10-08
34517

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 7/3/08 TIME 10:16
Intoxilyzer Serial # 80-002261 Location DUT HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: CMI INC
Serial Number: 80-002261
07/03/2008

10:16:03

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:16:31
Cal Check	0.101	10:16:50
Air Blank	0.000	10:17:19

Sarah Cione

Operator's Signature

7 of 7sc

Ch