



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
12-22-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12/12/08 Cal 1 - AC Cal 2 - AC

Between the dates of N/A and N/A the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 12/2/08.
The instrument was returned to service on _____.

Sarah Cione

Sarah L. Cione


12/22/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 22, day of December, 2008.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2 sc

INTOXILYZER SERVICE RECORD

T54
12-11-08

Date: 12/2/08

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-002261

Reason For Service: Breath hose will not heat up. Removed breath hose + tested it w/ breath hose testing device. Hose is good. Tried other hoses as well, still does not heat up.

Date Out of Service: 12/2/08

Calibration Check completed prior to removal from service

Date: -

SQAP completed prior to removal from service

Date: -

Corrective Action:

Sent to CMI, Inc.

Data was downloaded, but unable to complete Cal Check or SQAP.

2 of 2 sc



Return Material Authorization Form

This form **MUST** be completed and enclosed with item(s) being returned.
Failure to complete and return this form may cause delays in crediting / repairing.

1. Item(s) Returned: Instrument Supplies Other
(Note: please ship items in their original shipping container(s) or a similar protective box.)

2. Instrument Model: Intox 8000 Serial Number 80-002261

3. Bill To Address: Finance / Accounts Payable Ship To Address: (NO PO BOXES)
P.O. Box 27450 Tucson Police Department
TUCSON, AZ 85726-7450 Crime Lab
270 S. Stone Ave
TUCSON, AZ 85701-1917

Customer # 105075
Name Terry Gallegos
Phone: (520) 791-4494 Fax: (520) 791-4642
Email: Terry.Gallegos@tucsonaz.gov

4. Reason for Return: Note - If you are sending an item for repair, please give a detailed description of the problem. Please list any special instructions that you may have concerning this return. Note - Returns for credit may be subject to a 15% restocking charge.
Breath hose will not heat up. Tested breath hose w/ breath hose testing device, + hose is good. Tried other hoses, but still does not heat up.

5. Purchase Order Number (attach a copy of P.O. if applicable) 20669

6. I Authorize All Repairs: Yes No

7. I Authorize Repairs Up To: \$ _____

8. I Require An Estimate Regardless of Cost Yes No

Note: An evaluation fee will apply to estimates that are not repaired.

9. Authorized By: Terry Gallegos
Name (Please Print)
Terry Gallegos
Signature 12-11-2008
Date