CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

Dy 7.29.09

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 07/23/09 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 06/25/09 and 07/23/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Cheryl K. Straub

Cheryl K. Straub

07/28/09

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 28, day of July, 2009.

AMHernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

179 cs

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

PG 7-29-09

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 07/23/09 TIME 07:21

INTOXILYZER SERIAL # 80-002261 LOCATION TEAM II / DUI

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS "Sequence Aborted Test" on page 2 was stopped (Aborted) to first preform Exhibit G-2 cs

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

239
cs

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
07/23/2009 07:17:39

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:19:20
Diagnostic Test	Pass	07:19:47
Air Blank	ABT*	07:19:55
Air Blank	0.000	07:20:23

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

379
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
07/23/2009 07:23:16

B4

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:24:24
Diagnostic Test	Pass	07:24:52
Air Blank	0.000	07:25:19
0.100 Cal Check	0.100	07:25:38
Air Blank	0.000	07:26:08
Subject Test	0.000	07:26:38
Air Blank	ABT*	07:26:55
Air Blank	0.000	07:27:24

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

499
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
07/23/2009 07:27:46

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	07:28:39
Air Blank	IPS**	07:28:53

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

579
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
07/27/2009 11:59:40

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:00:50
Diagnostic Test	Pass	12:01:17
Air Blank	0.000	12:01:45
0.100 Cal Check	0.100	12:02:03
Air Blank	0.000	12:02:33
Subject Test	0.210	12:03:00
Air Blank	0.000	12:03:32
Five Minute Wait		
Air Blank	0.000	12:08:28
Subject Test	0.079*	12:11:37
Air Blank	0.000	12:12:09

*Deficient Sample

Blew too long for Deficient sample, ~~instrument~~ CS
instrument took it as Subject Test-
Test Repeated CS.

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

699
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
07/23/2009 07:29:14

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: ;. STRAUB #50772
TPD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:31:06
Diagnostic Test	Pass	07:31:33
Air Blank	0.000	07:32:01
0.100 Cal Check	0.100	07:32:19
Air Blank	0.000	07:32:49
Subject Test	0.073	07:34:54
Air Blank	0.000	07:35:27
Five Minute Wait		
Air Blank	0.000	07:40:22
Subject Test	XXX*	07:41:09
Air Blank	0.000	07:41:42

*Invalid Sample
Begin new deprivation period

while performing invalid sample test, ran out of air; instrument took it as subject test. Test repeated. CS

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

779
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
07/23/2009 07:42:36

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:43:03
Cal Check	0.099	07:43:22
Air Blank	0.000	07:43:52

754

Cheryl Straub
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

899
CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSH 7-28-09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department

DATE 07/23/09 TIME 07:21

Intoxilyzer Serial # 80-002261 Location TEAM II/DUI

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

999
CS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
07/23/2009 07:21:30

154

Standard Lot#: 808101G
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:21:58
Cal Check	0.100	07:22:16
Air Blank	0.000	07:22:46

Cheryl Straub
Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.