



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSY
4-12-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/1/10 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 3/4/10 and 4/1/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -
The instrument was returned to service on -

Sarah Cione

Sarah L. Cione

4/12/10

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 12th, day of April, 2010.

Amy Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

TS4
4.12.2010

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 4/1/10 TIME 924

INTOXILYZER SERIAL # 80-002261 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0.000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
04/01/2010

09:25:55

2 of 7 sc

104

Standard Lot#: 826001I
Last Changed By: C. NOLAN #38874

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:26:43
Diagnostic Test	Pass	09:27:10
Air Blank	0.000	09:27:37
0.100 Cal Check	0.100	09:27:56
Air Blank	0.000	09:28:26
Subject Test	0.000	09:28:45
Air Blank	0.000	09:29:12
Five Minute Wait	ABT*	
Air Blank	0.000	09:29:49

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
04/01/2010

09:30:03

3 of 7 sc

B4

Standard Lot#: 826001I
Last Changed By: C. NOLAN #38874

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:30:33
Air Blank	IPS**	09:30:41

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

4 of 7 sc

SA

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
04/01/2010 09:30:48

Standard Lot#: 826001I
Last Changed By: C. NOLAN #38874

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:31:34
Diagnostic Test	Pass	09:32:01
Air Blank	0.000	09:32:29
0.100 Cal Check	0.100	09:32:47
Air Blank	0.000	09:33:17
Subject Test	0.081*	09:36:25
Air Blank	0.000	09:36:57
Air Blank	0.000	09:37:24
Subject Test	XXX**	09:37:57
Air Blank	0.000	09:38:30

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
04/01/2010 09:39:33

5 of 7 sc

BSH

Standard Lot#: 826001I
Last Changed By: C. NOLAN #38874

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:40:00
Cal Check	0.099	09:40:19
Air Blank	0.000	09:40:48

Sarah Cione
Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TS4
4.12.2010

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 4/1/10 TIME 9:24
Intoxilyzer Serial # 80-002261 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
04/01/2010 09:24:02

7 of 7 sc

Standard Lot#: 826001I
Last Changed By: C. NOLAN #38874

124

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:24:29
Cal Check	0.099	09:24:48
Air Blank	0.000	09:25:17

Sarah Cione
Operator's Signature