TUCSON POLICE DEPARTMENT

CRIME LABORATORY 154

The undersigned Quality Assurance Specialist, having been qualified by the Arizona Department of Public Safety for Intoxilyzer 5000 inspection, and having performed such inspections on Intoxilyzer 5000 # 12, Serial # 64-002516, certifies by her signature:

on 1/26/04, with a result of .078 AC, 0.080 AC,

and on 2/19/04, with a result of .073 AC, 0.075 AC,

this instrument was tested with a known alcohol reference standard, and always produced results within the acceptable accuracy limits (from 0.070 AC. to 0.090 AC), as defined by R9-14-404.A.3.

No problems were reported concerning this instrument between these two dates.

Irene Doran

Irene Doran

ADPS QAS Permit #QAS 29586 (issue date 01/08/2004)

SUBSCRIBED AND SWORN to before me

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC - ARIZONA
PIMA COUNTY
My Comm. Expires Sept. 4, 2004

this 20, day of February, 2004

Kathy A. Kinney

Notary Public

P 186

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran / R. Misse (Print Name)

INTOXILYZER SERIAL # 64-002516 (#12) LOCATION CDS

DATE 1/26 ~~19~~ 2004 TIME 0934

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .348.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ * Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.~~078~~ 0.080 AC
Results: 0.078 AC 0.080 IED 0.080 AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: * moved clock forward 3 min.

SIGNATURE Irene Doran Rhonda M Misse

0. N/A ✓ alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
01/26/04 #12

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:49
CAL. CHECK	.078	09:49
AIR BLANK	.000	09:50

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: OPS

OPERATOR: DORAN 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516 #12
01/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:51
CAL. CHECK	.080	09:51
AIR BLANK	.000	09:51

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: OPS

OPERATOR: DORAN 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

P 396

P-596

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/20/04 Time 0934
Intoxilyzer Serial# 64-002514(14) Location CDS
QA Specialist Irene Doran / Missy
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(✓) 6. Push Start Test button or press enter on keyboard.

(✓) 7. Insert card in response to display.

(✓) 8. Air blank completed.

(✓) 9. Calibration check completed.
Test results 0.078 AC / 0.080 AC

(✓) 10. Air blank completed.

(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran Rhonda V Nuse

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

186
DS

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES 184
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran (40331)
(Print Name)
INTOXILYZER SERIAL #(12) 64-002516 LOCATION Tram 1 (ODS)
DATE February 19 ~~19~~ 2004 TIME 0904

DS Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

DS DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .353.

DS Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

DS Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

DS* Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

DS Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

DS Alcohol free subject Test result 0.000 AC

DS Error recognition logic system functioning
✓Invalid test printed

DS Proper sample recognition system
✓Invalid sample printed DS
✓Deficient sample printed DS

N/A Completeness of sample purge with collection tube**

DS Calibration standard 0.080 AC
Results: 0.073AC + 0.075AC ✓

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: * moved clock ahead 2 minutes - alcohol free subject, printer test check had already been completed

SIGNATURE Irene Doran

O.N/A DS alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516
02/19/04 #12

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:09
CAL. CHECK	.073	09:09
AIR BLANK	.000	09:10

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: ODS

OPERATOR: DOR AN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002516 #12
02/19/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	09:10
CAL. CHECK	.075	09:11
AIR BLANK	.000	09:11

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: ODS

OPERATOR: DOR AN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

396

596

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 2/19/04 Time 0904
Intoxilyzer Serial# (12) 04-002516 Location Tram 1 ODS
QA Specialist Irene Doran (46331)
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.073 + 0.075 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)