TUCSON POLICE DEPARTMENT

CRIME LABORATORY 164

The undersigned Quality Assurance Specialist, having been qualified by the Arizona Department of Public Safety for Intoxilyzer 5000 inspection, and having performed such inspections on Intoxilyzer 5000 # —, Serial # 66-001759, certifies by her signature:

on 1/26/04, with a result of 0.080 AC, 0.084 AC,

and on 2/19/04, with a result of 0.071 AC, 0.073 AC,

this instrument was tested with a known alcohol reference standard, and always produced results within the acceptable accuracy limits (from 0.070 AC, to 0.090 AC), as defined by R9-14-404.A.3.

No problems were reported concerning this instrument between these two dates.

Irene Doran

Irene Doran

ADPS QAS Permit #QAS 29586 (issue date 01/08/2004)

SUBSCRIBED AND SWORN to before me

this 20, day of February, 2004.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC - ARIZONA
PIMA COUNTY
My Comm. Expires Sept. 4, 2004

Kathy A Kinney

Notary Public

1 of 6
JS

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES BH
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran / Misse
(Print Name)

INTOXILYZER SERIAL #66-001759 LOCATION ODD
DATE Jan 26 ~~19~~ 2004 TIME 1433
JS (Actual time) instrument adjusted to match.

JS Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

JS DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .389.

JS Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

JS Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

JS* Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

JS Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

JS Alcohol free subject Test result 0.000 AC

JS Error recognition logic system functioning
Invalid test printed

JS Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

JS Calibration standard 0.080 AC
Results: 0.080AC / 0.084AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: * moved clock ahead four minutes. The alcohol free subject test was run prior to instrument clock adjustment.

SIGNATURE Irene Doran [signature]

0.NA alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER /
AZ MODEL 5000 SN 66-001759
01/26/04 14:45 /

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:44
CAL. CHECK	.080	14:45
AIR BLANK	.000	14:45

SUBJECT'S NAME

TIME FIRST OBSERVED

ODD
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER /
AZ MODEL 5000 SN 66-001759
01/26/04 14:47 /

TEST	BRAC VALUE	TIME
AIR BLANK	.000	14:46
CAL. CHECK	.084	14:46
AIR BLANK	.000	14:46

SUBJECT'S NAME

TIME FIRST OBSERVED

ODD
INSTRUMENT LOCATION

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

3 of 6

5 of 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/24/04 Time 1433 *
Intoxilyzer Serial# 66-PP1759 Location ODD
QA Specialist Irene Doran / Misse
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.080 AC /0.084 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran [signature]

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

* Actual time. Instrument adjusted to match.

186
LD

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* BH

AGENCY Tucson Police QA SPECIALIST Irene Doran (46331)
(Print Name)
INTOXILYZER SERIAL # 66-001759 LOCATION ODD
DATE February 19 ~~19~~ 2004 TIME 1543
LD

LD Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

LD DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .349.

LD Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

LD Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

LD Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

LD Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

LD Alcohol free subject Test result 0.000 AC

LD Error recognition logic system functioning
Invalid test printed

LD Proper sample recognition system
Invalid sample printed LD
Deficient sample printed LD

N/A Completeness of sample purge with collection tube**

LD Calibration standard 0.080 AC
Results: 0.071 AC + 0.073 AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran

0. N/A LD alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
02/19/04 15:52

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:51
CAL. CHECK	.071	15:51
AIR BLANK	.000	15:52

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: ODD

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-001759
02/19/04 15:53

TEST	BRAC VALUE	TIME
AIR BLANK	.000	15:53
CAL. CHECK	.073	15:53
AIR BLANK	.000	15:53

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: ODD

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

586

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 2/19/04 Time 1543
Intoxilyzer Serial# 66-001759 Location ODD
QA Specialist Irene Doran (46331)
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.071AC + 0.073AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)