TUCSON POLICE DEPARTMENT

CRIME LABORATORY BH

The undersigned Quality Assurance Specialist, having been qualified by the Arizona Department of Public Safety for Intoxilyzer 5000 inspection, and having performed such inspections on Intoxilyzer 5000 # — . Serial # 66-004798 . certifies by her signature:

on 1/26/04 , with a result of 0.077 AC. 0.078 AC.

and on 2/19/04 . with a result of 0. 079 AC. 0.078 AC.

this instrument was tested with a known alcohol reference standard, and always produced results within the acceptable accuracy limits (from 0.070 AC. to 0.090 AC) . as defined by R9-14-404.A.3.

No problems were reported concerning this instrument between these two dates.

Irene Doran

Irene Doran

ADPS QAS Permit #QAS 29586 (issue date 01/08/2004)

SUBSCRIBED AND SWORN to before me

this 20 , day of February , 2004

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC - ARIZONA
PIMA COUNTY
My Comm. Expires Sept. 4, 2004

Kathy A Kinney

Notary Public

106

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran / Nisse
(Print Name)

INTOXILYZER SERIAL # 66-004798 LOCATION OPW

DATE Jan 26 ~~19~~ 2004 TIME 1324

Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .334.

Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

Alcohol free subject Test result 0.000 AC

Error recognition logic system functioning
Invalid test printed

Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

Calibration standard 0.080 AC

Results: 0.077AC / 0.078 AC

Instrument operating properly and accurately. YES / NO

COMMENTS:

SIGNATURE Irene Doran Rhonda M Nisse

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
01/26/04 13:35

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:34
CAL. CHECK	.077	13:35
AIR BLANK	.000	13:35

SUBJECT'S NAME

ODW
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
01/26/04 13:37

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:36
CAL. CHECK	.073	13:36
AIR BLANK	.000	13:37

SUBJECT'S NAME

ODW
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

386

5 of 6

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

154

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/26/04 Time 1324
Intoxilyzer Serial# 66-004798 Location DDW
QA Specialist Irene Doran / Misk
(Print Name)

1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
   Standard value: 0.080 AC
2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
4. Attach simulator to the simulator entrance port on the Intoxilyzer.
5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
6. Push Start Test button or press enter on keyboard.
7. Insert card in response to display.
8. Air blank completed.
9. Calibration check completed.
   Test results 0.077 AC / 0.078 AC
10. Air blank completed.
11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran Rhonda M Nine

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

106
LD

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* 154

AGENCY Tucson Police QA SPECIALIST Irene Doran
(Print Name)
INTOXILYZER SERIAL # 66-004798 LOCATION ODW
DATE February 19 ~~19~~ 2004 TIME 1254
LD

LD Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

LD DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .346.

LD Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

LD Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

LD * Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

LD Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

LD Alcohol free subject Test result 0.000 AC

LD Error recognition logic system functioning
Invalid test printed

LD Proper sample recognition system
Invalid sample printed LD
Deficient sample printed LD

N/A Completeness of sample purge with collection tube**

LD Calibration standard 0.080 AC
Results: 0.079AC & 0.078AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: * Adjusted clock 3 min to coincide with back cell phone time

SIGNATURE Irene Doran

o. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
02/19/04 13:03

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:03
CAL. CHECK	.079	13:03
AIR BLANK	.000	13:03

SUBJECT'S NAME

ODW

TIME FIRST OBSERVED INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

TUCSON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 66-004798
02/19/04 13:05

TEST	BRAC VALUE	TIME
AIR BLANK	.000	13:04
CAL. CHECK	.078	13:04
AIR BLANK	.000	13:05

SUBJECT'S NAME

ODW

TIME FIRST OBSERVED INSTRUMENT LOCATION

DORAN (46331)

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

386

5 of 6
PD

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE TSH

Agency Tucson Police Department Date 2/19/04 Time 1254
Intoxilyzer Serial# 66-004798 Location ODW
QA Specialist Irene Doran (46331)
(Print Name)

(PD) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(PD) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(PD) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(PD) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

PD (PD) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

PD (PD) 6. Push Start Test button or press enter on keyboard.

PD (PD) 7. Insert card in response to display.

PD (PD) 8. Air blank completed.

PD (PD) 9. Calibration check completed.
Test results 0.079 + 0.078 AC

PD (PD) 10. Air blank completed.

PD (PD) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(PD) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)